

LB124410


7471B

INSTRUMENT ID: CVA

Linear

 μ Abs.:

64249



A= 0.0000e+000

B= 1.5599e-004

C= -4.3301e-002

Rho= 0.9999819

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	✓/D
0.0	0.000	-0.020	-0.020	149	0.000	149					-
0.2	0.200	0.220	0.020	1689	0.0 %	1689					10
2.5	2.500	2.476	-0.024	16149	0.0 %	16149					-1
5.0	5.000	5.030	0.030	32522	0.0 %	32522					1
7.5	7.500	7.515	0.015	48455	0.0 %	48455					0
10.0	10.000	9.979	-0.021	64249	0.0 %	64249					0

LB124410 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Mar 2023 15:08:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	149	-	0.0000	7471B	PPB	08 Mar 2023 15:13:54	S	Std
0.2 - 1	50.2	1689	-	0.2000	7471B	PPB	08 Mar 2023 15:16:12	S	Std
2.5 - 1	52.5	16149	-	2.5000	7471B	PPB	08 Mar 2023 15:18:29	S	Std
5.0 - 1	55	32522	-	5.0000	7471B	PPB	08 Mar 2023 15:20:46	S	Std
7.5 - 1	57.5	48455	-	7.5000	7471B	PPB	08 Mar 2023 15:23:03	S	Std
10.0 - 1	58	64249	-	10.0000	7471B	PPB	08 Mar 2023 15:25:21	S	Std
ICV75 - 1	ICV75	28062	4.3341	-	7471B	PPB	08 Mar 2023 15:28:02	U	SMPL
ICB75 - 1	ICB75	61	-0.0338	-	7471B	PPB	08 Mar 2023 15:30:17	U	SMPL
CCV38 - 1	CCV38	31440	4.8611	-	7471B	PPB	08 Mar 2023 15:32:33	U	SMPL
CCB38 - 1	CCB38	66	-0.0330	-	7471B	PPB	08 Mar 2023 15:34:49	U	SMPL
CRA - 1	CRA	1687	0.2199	-	7471B	PPB	08 Mar 2023 15:37:05	U	SMPL
HighStd - 1	HighStd	64166	9.9660	-	7471B	PPB	08 Mar 2023 15:39:20	U	SMPL
ChkStd - 1	ChkStd	45543	7.0610	-	7471B	PPB	08 Mar 2023 15:41:36	U	SMPL
PB151318BL - 1	PBS	-60	-0.0527	-	7471B	PPB	08 Mar 2023 15:43:54	U	SMPL
PB151318BS - 1	LCSS	23617	3.6407	-	7471B	PPB	08 Mar 2023 15:46:12	U	SMPL
O1821-01 - 1	WC-IDW-S-1	9934	1.5063	-	7471B	PPB	08 Mar 2023 15:48:29	U	SMPL
O1825-01 - 1	HH-TP1-0306	2947	0.4164	-	7471B	PPB	08 Mar 2023 15:50:46	U	SMPL
O1825-03 - 1	HH-TP2-0306	2551	0.3546	-	7471B	PPB	08 Mar 2023 15:53:03	U	SMPL
O1825-05 - 1	HH-TP3-0306	4332	0.6325	-	7471B	PPB	08 Mar 2023 15:55:20	U	SMPL
O1825-07 - 1	HH-TP4-0306	3431	0.4919	-	7471B	PPB	08 Mar 2023 15:57:36	U	SMPL
CCV39 - 1	CCV39	31641	4.8924	-	7471B	PPB	08 Mar 2023 15:59:51	U	SMPL
CCB39 - 1	CCB39	78	-0.0311	-	7471B	PPB	08 Mar 2023 16:02:07	U	SMPL
O1827-01 - 1	EO-03-030623	3014	0.4269	-	7471B	PPB	08 Mar 2023 16:04:23	U	SMPL
O1828-01 - 1	AVENEL-60	6576	0.9825	-	7471B	PPB	08 Mar 2023 16:06:39	U	SMPL
O1833-01 - 1	46	27302	4.2156	-	7471B	PPB	08 Mar 2023 16:08:55	U	SMPL
O1833-04 - 1	42	57372	8.9062	-	7471B	PPB	08 Mar 2023 16:11:11	U	SMPL
O1834-01 - 1	TP-7	966	0.1074	-	7471B	PPB	08 Mar 2023 16:25:17	U	SMPL
O1834-02 - 1	TP-8	3938	0.5710	-	7471B	PPB	08 Mar 2023 16:27:34	U	SMPL
CCV40 - 1	CCV40	31044	4.7993	-	7471B	PPB	08 Mar 2023 16:29:49	U	SMPL
CCB40 - 1	CCB40	78	-0.0311	-	7471B	PPB	08 Mar 2023 16:32:05	U	SMPL
O1834-03 - 1	TP-9	858	0.0905	-	7471B	PPB	08 Mar 2023 16:34:22	U	SMPL
O1834-04 - 1	TP-10	2133	0.2894	-	7471B	PPB	08 Mar 2023 16:36:37	U	SMPL
PB151319BL - 1	PBS	177	-0.0157	-	7471B	PPB	08 Mar 2023 16:38:53	U	SMPL
PB151319BS - 1	LCSS	23523	3.6261	-	7471B	PPB	08 Mar 2023 16:41:09	U	SMPL
O1829-02 - 1	SS-02-20230306	1631	0.2111	-	7471B	PPB	08 Mar 2023 16:55:20	U	SMPL
O1829-03 - 1	SS-03-20230306	674	0.0618	-	7471B	PPB	08 Mar 2023 16:57:37	U	SMPL
GGV41 - 1	CCV41	937	0.1029	-	7471B	PPB	08 Mar 2023 16:59:52	U	SMPL
CCB41 - 1	CCB41	0	-0.0433	-	7471B	PPB	08 Mar 2023 17:02:08	U	SMPL
CCV41 - 1	CCV41	30050	4.6442	-	7471B	PPB	08 Mar 2023 17:09:11	U	SMPL
CCB41 - 1	CCB41	8	-0.0421	-	7471B	PPB	08 Mar 2023 17:11:27	U	SMPL
O1829-04 - 1	SS-04-20230306	1244	0.1508	-	7471B	PPB	08 Mar 2023 17:13:45	U	SMPL
O1829-05 - 1	SS-05-20230306	3635	0.5237	-	7471B	PPB	08 Mar 2023 17:16:00	U	SMPL
O1829-06 - 1	SS-06-20230306	940	0.1033	-	7471B	PPB	08 Mar 2023 17:18:16	U	SMPL
O1829-07 - 1	SS-07-20230306	1575	0.2024	-	7471B	PPB	08 Mar 2023 17:20:32	U	SMPL
O1829-08 - 1	SS-08-20230306	933	0.1022	-	7471B	PPB	08 Mar 2023 17:22:48	U	SMPL
O1829-09 - 1	SS-09-20230306	866	0.0918	-	7471B	PPB	08 Mar 2023 17:25:04	U	SMPL
O1829-10 - 1	SS-10-20230306	4084	0.5938	-	7471B	PPB	08 Mar 2023 17:27:20	U	SMPL
O1829-11 - 1	O1829-10MS	4471	0.6541	-	7471B	PPB	08 Mar 2023 17:29:38	U	SMPL
O1829-12 - 1	O1829-10MSD	4285	0.6251	-	7471B	PPB	08 Mar 2023 17:31:55	U	SMPL
O1829-13 - 1	SS-11-20230306	4311	0.6292	-	7471B	PPB	08 Mar 2023 17:34:11	U	SMPL
CCV42 - 1	CCV42	31750	4.9094	-	7471B	PPB	08 Mar 2023 17:36:27	U	SMPL
CCB42 - 1	CCB42	55	-0.0347	-	7471B	PPB	08 Mar 2023 17:38:42	U	SMPL

MS
3/10/23