



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124422

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/09/2023	11:05
2	CAL2	1	Water	NA	NA	20.7	7.00	03/09/2023	11:06
3	CAL3	1	Water	NA	NA	20.7	10.03	03/09/2023	11:08
4	ICV	1	Water	NA	NA	20.8	6.98	03/09/2023	11:10
5	CCV1	1	Water	NA	NA	20.8	2.01	03/09/2023	11:15
6	O1852-02	1	Water	NA	NA	20.3	5.95	03/09/2023	11:22
7	O1852-02DUP	1	Water	NA	NA	20.4	5.96	03/09/2023	11:23
8	O1852-03	1	Water	NA	NA	20.1	6.80	03/09/2023	11:25
9	CCV2	1	Water	NA	NA	20.8	12.02	03/09/2023	11:30

124422

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 01852 WorkList ID : 168032 Department : Wet-Chemistry Date : 03-09-2023 10:55:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/08/2023	Water	O1852-02	pH	Cool 4 deg C	TETR06	K13	RW5-IDWGW-20230308	03/08/2023	9040C
03/08/2023	Water	O1852-03	pH	Cool 4 deg C	TETR06	K13	TT-051-IDWGW-20230308	03/08/2023	9040C

Date/Time 03/09/23 11:00
Raw Sample Received by: ZP(WC)
Raw Sample Relinquished by: [Signature]

Date/Time 03/09/23 12:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: ZP(WC)