

LB 124434


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48500



A= 0.0000e+000

B= 2.0910e-004

C= -9.2240e-003

Rho= 0.9987631

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.056	0.056	314	0.000	314					15
0.2	0.200	0.230	0.030	1142	0.0 %	1142					-6
2.5	2.500	2.357	-0.143	11318	0.0 %	11318					5
5.0	5.000	5.243	0.243	25116	0.0 %	25116					-4
7.5	7.500	7.182	-0.318	34390	0.0 %	34390					1
10.0	10.000	10.132	0.132	48500	0.0 %	48500					

LB124434 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Mar 2023 09:43:23

Sample ID	Extended ID	u Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	314	-	0.0000	7470A	PPB	10 Mar 2023 09:50:02	S	Std
0.2 - 1	50.2	1142	-	0.2000	7470A	PPB	10 Mar 2023 09:52:20	S	Std
2.5 - 1	52.5	11318	-	2.5000	7470A	PPB	10 Mar 2023 09:54:37	S	Std
5.0 - 1	55	25116	-	5.0000	7470A	PPB	10 Mar 2023 10:02:37	S	Std
7.5 - 1	57.5	34390	-	7.5000	7470A	PPB	10 Mar 2023 10:04:54	S	Std
10.0 - 1	510	48500	-	10.0000	7470A	PPB	10 Mar 2023 10:09:45	S	Std
ICV78 - 1	ICV78	18167	3.7895	-	7470A	PPB	10 Mar 2023 10:12:29	U	SMPL
ICB78 - 1	ICB78	-52	-0.0201	-	7470A	PPB	10 Mar 2023 10:14:44	U	SMPL
CCV43 - 1	CCV43	25084	5.2359	-	7470A	PPB	10 Mar 2023 10:17:00	U	SMPL
CCB43 - 1	CCB43	-74	-0.0247	-	7470A	PPB	10 Mar 2023 10:19:15	U	SMPL
CRA - 1	CRA	932	0.1857	-	7470A	PPB	10 Mar 2023 10:21:31	U	SMPL
HighStd - 1	HighStd	48013	10.0304	-	7470A	PPB	10 Mar 2023 10:23:48	U	SMPL
ChkStd - 1	ChkStd	30616	6.3928	-	7470A	PPB	10 Mar 2023 10:26:03	U	SMPL
PB151344BL - 1	PBW	-88	-0.0276	-	7470A	PPB	10 Mar 2023 10:28:21	U	SMPL
PB151344BS - 1	LCSW	18057	3.7665	-	7470A	PPB	10 Mar 2023 10:33:11	U	SMPL
O1627-07 - 1	LOD-MDL-WATER-01-QT1-2023	274	0.0481	-	7470A	PPB	10 Mar 2023 10:38:07	U	SMPL
O1627-08 - 1	LOQ-WATER-02-QT1-2023	981	0.1959	-	7470A	PPB	10 Mar 2023 10:40:23	U	SMPL
O1832-01 - 1	B-19-GW-01	2442	0.5014	-	7470A	PPB	10 Mar 2023 10:42:40	U	SMPL
O1835-06 - 1	COMP	678	0.1325	-	7470A	PPB	10 Mar 2023 10:44:57	U	SMPL
O1852-02 - 1	RW5-IDWGW-20230308	36	-0.0017	-	7470A	PPB	10 Mar 2023 10:47:12	U	SMPL
CCV44 - 1	CCV44	24536	5.1213	-	7470A	PPB	10 Mar 2023 10:49:28	U	SMPL
CCB44 - 1	CCB44	-46	-0.0188	-	7470A	PPB	10 Mar 2023 10:51:43	U	SMPL
O1852-03 - 1	TT-051-IDWGW-20230308	64	0.0042	-	7470A	PPB	10 Mar 2023 10:54:00	U	SMPL
O1853-03 - 1	20230167-COMPOSITE-1-ELUTRIATE	77	0.0069	-	7470A	PPB	10 Mar 2023 10:56:15	U	SMPL
O1853-03DUP - 1	20230167-COMPOSITE-1-ELUTRIATEDUP	78	0.0071	-	7470A	PPB	10 Mar 2023 10:58:31	U	SMPL
O1853-03MS - 1	20230167-COMPOSITE-1-ELUTRIATEMS	18162	3.7885	-	7470A	PPB	10 Mar 2023 11:03:19	U	SMPL
O1853-03MSD - 1	20230167-COMPOSITE-1-ELUTRIATEMSD	17995	3.7536	-	7470A	PPB	10 Mar 2023 11:05:35	U	SMPL
O1853-04 - 1	20230168-COMPOSITE-2-ELUTRIATE	-108	-0.0318	-	7470A	PPB	10 Mar 2023 11:07:57	U	SMPL
O1853-03LX5 - 1		-21	-0.0136	-	7470A	PPB	10 Mar 2023 11:10:17	U	SMPL
O1853-03A - 1		18030	3.7609	-	7470A	PPB	10 Mar 2023 11:12:35	U	SMPL
CCV45 - 1	CCV45	24900	5.1974	-	7470A	PPB	10 Mar 2023 11:17:06	U	SMPL
CCB45 - 1	CCB45	-328	-0.0778	-	7470A	PPB	10 Mar 2023 11:19:20	U	SMPL