


63124487

SFAM01.1 INSTRUMENT ID: CV1

Linear

μ Abs.:

51084



A= 0.0000e+000

B= 1.9735e-004 slope

C= 2.2730e-002 y-intercept

Rho= 0.9996594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/bD
0.00	0.000	0.041	0.041	93	0.000	93					5
0.05	0.050					185	0				14
0.20	0.200	0.228	0.028	1040	0.0 %	1040					-3
2.50	2.500	2.421	-0.079	12154	0.0 %	12154					2
5.00	5.000	5.079	0.079	25620	0.0 %	25620					-2
7.50	7.500	7.327	-0.173	37012	0.0 %	37012					1
10.0	10.000	10.104	0.104	51084	0.0 %	51084					

LB124437

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	93	-		0 SFAM01.1	PPB	3/10/2023 12:08	Std
	0.2 S01	1040	-		0.2 SFAM01.1	PPB	3/10/2023 12:11	Std
	2.5 S02	12154	-		2.5 SFAM01.1	PPB	3/10/2023 12:13	Std
	5 S03	25620	-		5 SFAM01.1	PPB	3/10/2023 12:15	Std
	7.5 S04	37012	-		7.5 SFAM01.1	PPB	3/10/2023 12:18	Std
	10 S05	51084	-		10 SFAM01.1	PPB	3/10/2023 12:22	Std
ICV079	ICV079	21854	4.3355 -		SFAM01.1	PPB	3/10/2023 12:25	SMPL
ICB079	ICB079	22	0.0271 -		SFAM01.1	PPB	3/10/2023 12:27	SMPL
CCV046	CCV046	25289	5.0134 -		SFAM01.1	PPB	3/10/2023 12:36	SMPL
CCB046	CCB046	28	0.0283 -		SFAM01.1	PPB	3/10/2023 12:38	SMPL
PB151356BL	PBW356	89	0.0403 -		SFAM01.1	PPB	3/10/2023 12:40	SMPL
O1734-01	YBKW1	552	0.1317 -		SFAM01.1	PPB	3/10/2023 12:42	SMPL
O1734-02	YBKW3	1082	0.2363 -		SFAM01.1	PPB	3/10/2023 12:45	SMPL
O1734-03	YBKW5	1372	0.2935 -		SFAM01.1	PPB	3/10/2023 12:47	SMPL
O1734-04	YBKW7	1489	0.3166 -		SFAM01.1	PPB	3/10/2023 12:49	SMPL
O1734-05	YBKX3	266	0.0752 -		SFAM01.1	PPB	3/10/2023 12:51	SMPL
O1734-06	YBKX5	807	0.182 -		SFAM01.1	PPB	3/10/2023 12:54	SMPL
O1734-07	YBKX7	828	0.1861 -		SFAM01.1	PPB	3/10/2023 12:56	SMPL
O1734-08	YBKX9	382	0.0981 -		SFAM01.1	PPB	3/10/2023 12:58	SMPL
O1734-09	YBKY1	519	0.1252 -		SFAM01.1	PPB	3/10/2023 13:00	SMPL
O1734-10	YBKY5	223	0.0667 -		SFAM01.1	PPB	3/10/2023 13:03	SMPL
O1734-11	YBKZ1	232	0.0685 -		SFAM01.1	PPB	3/10/2023 13:05	SMPL
O1734-12	YBL03	521	0.1255 -		SFAM01.1	PPB	3/10/2023 13:07	SMPL
O1734-13	YBL05	1043	0.2286 -		SFAM01.1	PPB	3/10/2023 13:09	SMPL
O1734-14	YBL07	297	0.0813 -		SFAM01.1	PPB	3/10/2023 13:12	SMPL
O1734-15	YBL07D	281	0.0782 -		SFAM01.1	PPB	3/10/2023 13:14	SMPL
O1734-16	YBL07S	4545	0.9197 -		SFAM01.1	PPB	3/10/2023 13:16	SMPL
O1734-17	YBL11	454	0.1123 -		SFAM01.1	PPB	3/10/2023 13:19	SMPL
O1734-18	YBL13	277	0.0774 -		SFAM01.1	PPB	3/10/2023 13:21	SMPL
O1734-19	YBL15	98	0.0421 -		SFAM01.1	PPB	3/10/2023 13:23	SMPL
CCV047	CCV047	24660	4.8893 -		SFAM01.1	PPB	3/10/2023 13:25	SMPL
CCB047	CCB047	39	0.0304 -		SFAM01.1	PPB	3/10/2023 13:28	SMPL
O1734-20	YBL32	253	0.0727 -		SFAM01.1	PPB	3/10/2023 13:30	SMPL
O1734-21	YBL53	109	0.0442 -		SFAM01.1	PPB	3/10/2023 13:32	SMPL
PB151357BL	PBW357	40	0.0306 -		SFAM01.1	PPB	3/10/2023 13:34	SMPL
O1771-20	MCOAY2	28	0.0283 -		SFAM01.1	PPB	3/10/2023 13:37	SMPL
PB151358BL	PBW358	128	0.048 -		SFAM01.1	PPB	3/10/2023 13:39	SMPL
O1378-01	MDL-WATER	192	0.0606 -		SFAM01.1	PPB	3/10/2023 13:44	SMPL
CCV048	CCV048	24958	4.9481 -		SFAM01.1	PPB	3/10/2023 13:46	SMPL
CCB048	CCB048	33	0.0292 -		SFAM01.1	PPB	3/10/2023 13:48	SMPL