



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124438

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/10/2023	08:40
2	O1854-01		1	14	100.00	>212.0	>212.0	03/10/2023	09:10
3	O1854-02		1	13	100.00	>212.0	>212.0	03/10/2023	09:40
4	O1854-02DUP		1	14	100.00	>212.0	>212.0	03/10/2023	10:10
5	O1854-03		1	14	100.00	>212.0	>212.0	03/10/2023	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB124438

WorkList Name : fp-3-9

WorkList ID : 168042

Department : Wet-Chemistry

Date : 03-09-2023 15:31:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/16/2023	Water	O1854-01	Flash Point	Cool 4 deg C	PSEG03	K13	31844	03/09/2023	1010B
03/16/2023	Water	O1854-02	Flash Point	Cool 4 deg C	PSEG03	K13	33907	03/09/2023	1010B
03/16/2023	Water	O1854-03	Flash Point	Cool 4 deg C	PSEG03	K13	33952	03/09/2023	1010B

Date/Time 03/10/2023 08:30
 Raw Sample Received by: Rm
 Raw Sample Relinquished by: J (WC)

Date/Time 03/10/2023 11:20
 Raw Sample Received by: J (WC)
 Raw Sample Relinquished by: Rm