

LB124471


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

57376



A= 0.0000e+000

B= 1.7432e-004

C= -8.3584e-002

Rho= 0.9997591

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	YLD
0.0	0.000	0.024	0.024	616	0.000	616					
0.2	0.200	0.195	-0.005	1599	0.0 %	1599					-3
2.5	2.500	2.475	-0.025	14678	0.0 %	14678					-1
5.0	5.000	4.927	-0.073	28741	0.0 %	28741					-1
7.5	7.500	7.661	0.161	44430	0.0 %	44430					2
10.0	10.000	9.918	-0.082	57376	0.0 %	57376					-1

LB124471 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Mar 2023 16:10:47

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	616	-	0.0000	7471B	PPB	13 Mar 2023 16:14:24	S	Std
0.2 - 1	50.2	1599	-	0.2000	7471B	PPB	13 Mar 2023 16:16:41	S	Std
2.5 - 1	52.5	14678	-	2.5000	7471B	PPB	13 Mar 2023 16:18:58	S	Std
5.0 - 1	55	28741	-	5.0000	7471B	PPB	13 Mar 2023 16:21:16	S	Std
7.5 - 1	57.5	44430	-	7.5000	7471B	PPB	13 Mar 2023 16:28:52	S	Std
10.0 - 1	511	57376	-	10.0000	7471B	PPB	13 Mar 2023 16:31:08	S	Std
ICV83 - 1	ICV83	24726	4.2266	-	7471B	PPB	13 Mar 2023 16:34:45	U	SMPL
ICB83 - 1	ICB83	483	0.0006	-	7471B	PPB	13 Mar 2023 16:37:00	U	SMPL
CCV56 - 1	CCV56	28296	4.8489	-	7471B	PPB	13 Mar 2023 16:39:17	U	SMPL
CCB56 - 1	CCB56	485	0.0010	-	7471B	PPB	13 Mar 2023 16:41:32	U	SMPL
CRA - 1	CRA	1496	0.1772	-	7471B	PPB	13 Mar 2023 16:43:49	U	SMPL
HighStd - 1	HighStd	56836	9.8240	-	7471B	PPB	13 Mar 2023 16:46:04	U	SMPL
ChkStd - 1	ChkStd	42439	7.3143	-	7471B	PPB	13 Mar 2023 16:48:20	U	SMPL
PB151397BL - 1	PBS	383	-0.0168	-	7471B	PPB	13 Mar 2023 16:50:38	U	SMPL
PB151397BS - 1	LCSS	21419	3.6501	-	7471B	PPB	13 Mar 2023 16:56:09	U	SMPL
O1627-01 - 1	LOD-MDL-SOIL-01-QT1-2023	748	0.0468	-	7471B	PPB	13 Mar 2023 17:01:02	U	SMPL
O1627-02 - 1	LOQ-SOIL-02-QT1-2023	1489	0.1760	-	7471B	PPB	13 Mar 2023 17:03:18	U	SMPL
O1853-01 - 1	20230162-COMPOSITE-1	1113	0.1104	-	7471B	PPB	13 Mar 2023 17:05:36	U	SMPL
O1853-02 - 1	20230163-COMPOSITE-2	863	0.0669	-	7471B	PPB	13 Mar 2023 17:07:53	U	SMPL
O1857-01 - 1	RT-70-1	2205	0.3008	-	7471B	PPB	13 Mar 2023 17:10:09	U	SMPL
CCV57 - 1	CCV57	27839	4.7693	-	7471B	PPB	13 Mar 2023 17:12:25	U	SMPL
CCB57 - 1	CCB57	489	0.0017	-	7471B	PPB	13 Mar 2023 17:14:40	U	SMPL
O1876-01 - 1	38414-CONCRETE	1679	0.2091	-	7471B	PPB	13 Mar 2023 17:16:57	U	SMPL
O1877-01 - 1	COMP-SOIL	6124	0.9839	-	7471B	PPB	13 Mar 2023 17:25:05	U	SMPL
O1877-01DUP - 1	COMP-SOILDUP	6070	0.9745	-	7471B	PPB	13 Mar 2023 17:29:47	U	SMPL
O1877-01MS - 1	COMP-SOILMS	27755	4.7546	-	7471B	PPB	13 Mar 2023 17:34:43	U	SMPL
O1877-01MSD - 1	COMP-SOILMSD	27624	4.7318	-	7471B	PPB	13 Mar 2023 17:36:59	U	SMPL
O1877-01LX5 - 1		1466	0.1720	-	7471B	PPB	13 Mar 2023 17:41:54	U	SMPL
O1877-01A - 1		27422	4.6966	-	7471B	PPB	13 Mar 2023 17:44:10	U	SMPL
CCV58 - 1	CCV58	27887	4.7776	-	7471B	PPB	13 Mar 2023 17:46:27	U	SMPL
CCB58 - 1	CCB58	788	0.0538	-	7471B	PPB	13 Mar 2023 17:48:45	U	SMPL