



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124480

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP101471
Turbidity Calibration std, 0NTU	WP101470
Turbidity Calibration std, 20NTU	WP101463
Turbidity Calibration std, 40NTU	WP101461
Turbidity Calibration std, 1NTU	WP101467
Turbidity Calibration std, 5NTU	WP101466
Turbidity LOQ std, 1.0NTU	WP101472
Turbidity Calibration - CCV std, 10 NTU	WP101464
10 NTU Standard 500 ml	W2971

Intercept: 0.0075

Slope: 0.9328

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.15		03/14/2023	09:20
2	CAL2	1	1	1.150	1.23	22.5	03/14/2023	09:24
3	CAL3	5	1	4.380	4.69	-6.2	03/14/2023	09:27
4	CAL4	10	1	9.460	10.13	1.3	03/14/2023	09:30
5	CAL5	20	1	18.300	19.61	-2	03/14/2023	09:33
6	CAL6	40	1	37.500	40.19	0.5	03/14/2023	09:37



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124480

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.300	9.962	03/14/2023	09:40
2	ICB		1	0.138	0.140	03/14/2023	09:44
3	CCV1	10	1	9.360	10.026	03/14/2023	09:47
4	CCB1		1	0.148	0.151	03/14/2023	09:50
5	LB124480BL		1	0.152	0.155	03/14/2023	09:53
6	O1627-07		1	0.492	0.519	03/14/2023	09:57
7	O1627-08		1	1.116	1.188	03/14/2023	10:00
8	CCV2	10	1	9.180	9.833	03/14/2023	10:03
9	CCB2		1	0.150	0.153	03/14/2023	10:07