



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124481

| Reagent/Standard                        | Lot/Log # |
|-----------------------------------------|-----------|
| TURBIDITY LOD STD, 0.5NTU               | WP101471  |
| Turbidity Calibration std, 0NTU         | WP101470  |
| Turbidity Calibration std, 20NTU        | WP101463  |
| Turbidity Calibration std, 40NTU        | WP101461  |
| Turbidity Calibration std, 1NTU         | WP101467  |
| Turbidity Calibration std, 5NTU         | WP101466  |
| Turbidity LOQ std, 1.0NTU               | WP101472  |
| Turbidity Calibration - CCV std, 10 NTU | WP101464  |
| 10 NTU Standard 500 ml                  | W2971     |

Intercept: 0.0075

Slope: 0.9328

Regression: 0.9998

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Result (mg/l) | %D   | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|---------------|------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.146   | 0.15          |      | 03/14/2023 | 09:20     |
| 2   | CAL2   | 1                | 1        | 1.150   | 1.23          | 22.5 | 03/14/2023 | 09:24     |
| 3   | CAL3   | 5                | 1        | 4.380   | 4.69          | -6.2 | 03/14/2023 | 09:27     |
| 4   | CAL4   | 10               | 1        | 9.460   | 10.13         | 1.3  | 03/14/2023 | 09:30     |
| 5   | CAL5   | 20               | 1        | 18.300  | 19.61         | -2   | 03/14/2023 | 09:33     |
| 6   | CAL6   | 40               | 1        | 37.500  | 40.19         | 0.5  | 03/14/2023 | 09:37     |



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124481

| Seq | Lab ID     | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|------------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV        | 10                  | 1        | 9.300   | 9.962           | 03/14/2023 | 09:40     |
| 2   | ICB        |                     | 1        | 0.138   | 0.140           | 03/14/2023 | 09:44     |
| 3   | CCV1       | 10                  | 1        | 9.360   | 10.026          | 03/14/2023 | 09:47     |
| 4   | CCB1       |                     | 1        | 0.148   | 0.151           | 03/14/2023 | 09:50     |
| 5   | LB124481BL |                     | 1        | 0.152   | 0.155           | 03/14/2023 | 09:53     |
| 6   | O1627-10   |                     | 1        | 0.492   | 0.519           | 03/14/2023 | 09:57     |
| 7   | O1627-11   |                     | 1        | 1.116   | 1.188           | 03/14/2023 | 10:00     |
| 8   | CCV2       | 10                  | 1        | 9.180   | 9.833           | 03/14/2023 | 10:03     |
| 9   | CCB2       |                     | 1        | 0.150   | 0.153           | 03/14/2023 | 10:07     |