



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124484

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP101484
Residual chlorine ICV-LCS, 0.4PPM	WP101483
Chlorine Calibration std, 0.1ppm	WP101478
Chlorine Calibration std, 0.2ppm	WP101479
Chlorine Calibration std, 0.8ppm	WP101481
Chlorine Calibration std, 0.0ppm	WP101477
Chlorine Calibration std, 1.6ppm	WP101482
Residual Chlorine LOQ 0.1ppm	WP101485
Residual Chlorine Calibration and CCV std, 0	WP101480
Total Chlorine Powder Pillows	W2759

Intercept: 0.0149

Slope: 0.9971

Regression: 0.999800

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.02		03/14/2023	10:30
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	5	03/14/2023	10:33
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	-2	03/14/2023	10:37
4	CAL4	0.4	1	0.000	0.420	0.420	0.41	1.5	03/14/2023	10:40
5	CAL5	0.8	1	0.000	0.830	0.830	0.82	2.1	03/14/2023	10:43
6	CAL6	1.6	1	0.000	1.600	1.600	1.59	-0.6	03/14/2023	10:47

Analysis Method: SM4500 Cl GANALYST: IwonaParameter: Residual ChlorineSUPERVISOR REVIEW BY: SohilRun Number: LB124484

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.4160	03/14/2023	10:50
2	ICB				1	0.0000	0.0000	0.0000	-0.0150	03/14/2023	10:53
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4060	03/14/2023	10:56
4	CCB1				1	0.0000	0.0100	0.0100	-0.0050	03/14/2023	11:00
5	LB124484BL	50	50		1	0.0000	0.0100	0.0100	-0.0050	03/14/2023	11:03
6	LB124484BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3760	03/14/2023	11:07
7	O1627-07	50	50		1	0.0000	0.0600	0.0600	0.0450	03/14/2023	11:10
8	O1627-08	50	50		1	0.0000	0.1300	0.1300	0.1150	03/14/2023	11:14
9	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3960	03/14/2023	11:17
10	CCB2				1	0.0000	0.0000	0.0000	-0.0150	03/14/2023	11:20