



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124489

Slope : 98.6

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/14/2023	14:45
2	CAL2	1	Water	NA	NA	20.6	7.00	03/14/2023	14:46
3	CAL3	1	Water	NA	NA	20.6	10.03	03/14/2023	14:49
4	ICV	1	Water	NA	NA	20.7	6.98	03/14/2023	14:50
5	CCV1	1	Water	NA	NA	20.7	2.01	03/14/2023	14:55
6	O1907-08	1	Water	NA	NA	20.3	6.57	03/14/2023	15:00
7	O1910-01	1	Water	NA	NA	20.1	4.37	03/14/2023	15:01
8	O1910-02	1	Water	NA	NA	20.2	4.58	03/14/2023	15:05
9	O1910-03	1	Water	NA	NA	20.1	4.55	03/14/2023	15:10
10	CCV2	1	Water	NA	NA	20.7	12.03	03/14/2023	15:15

12489

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1907

WorkList ID : 168150

Department : Wet-Chemistry

Date : 03-14-2023 14:25:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/14/2023	Water	O1907-08	pH	Cool 4 deg C	PSEG03	J11	33969	03/14/2023	9040C
03/13/2023	Water	O1910-01	pH	Cool 4 deg C	STER07	J11	CELL-A	03/13/2023	9040C
03/13/2023	Water	O1910-02	pH	Cool 4 deg C	STER07	J11	CELL-B-C	03/13/2023	9040C
03/13/2023	Water	O1910-03	pH	Cool 4 deg C	STER07	J11	CELL-D	03/13/2023	9040C

Date/Time 03-14-23 14:35

Raw Sample Received by: J.P. (W.C.)

Raw Sample Relinquished by: J.P. (W.C.)

Date/Time 03-14-23 16:00

Raw Sample Received by: J.P. (W.C.)

Raw Sample Relinquished by: J.P. (W.C.)