

LB124502


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

44260



A= 0.0000e+000

B= 2.2464e-004

C= 1.7807e-003

Rho= 0.9998658

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/O
0.0	0.000	0.011	0.011	41	0.000	41					0.00
0.2	0.200	0.221	0.021	975	0.0 %	975					11
2.5	2.500	2.398	-0.102	10668	0.0 %	10668					4
5.0	5.000	5.063	0.063	22530	0.0 %	22530					1
7.5	7.500	7.563	0.063	33659	0.0 %	33659					1
10.0	10.000	9.944	-0.056	44260	0.0 %	44260					1

LB124502 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Mar 2023 11:14:00

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	15 Mar 2023 11:30:27	S	Std
0.2 - 1	50.2	975	-	0.2000	7470A	PPB	15 Mar 2023 11:32:45	S	Std
2.5 - 1	52.5	10668	-	2.5000	7470A	PPB	15 Mar 2023 11:35:02	S	Std
5.0 - 1	55	22530	-	5.0000	7470A	PPB	15 Mar 2023 11:37:19	S	Std
7.5 - 1	57.5	33659	-	7.5000	7470A	PPB	15 Mar 2023 11:39:36	S	Std
10.0 - 1	511	44260	-	10.0000	7470A	PPB	15 Mar 2023 11:41:52	S	Std
ICV86 - 1	ICV86	18472	4.1513	-	7470A	PPB	15 Mar 2023 11:44:43	U	SMPL
ICB86 - 1	ICB86	-28	-0.0045	-	7470A	PPB	15 Mar 2023 11:47:00	U	SMPL
CCV66 - 1	CCV66	22290	5.0090	-	7470A	PPB	15 Mar 2023 11:49:15	U	SMPL
CCB66 - 1	CCB66	-19	-0.0025	-	7470A	PPB	15 Mar 2023 11:51:30	U	SMPL
CRA - 1	CRA	954	0.2161	-	7470A	PPB	15 Mar 2023 11:53:45	U	SMPL
HighStd - 1	HighStd	43721	9.8232	-	7470A	PPB	15 Mar 2023 11:56:01	U	SMPL
ChkStd - 1	ChkStd	31210	7.0127	-	7470A	PPB	15 Mar 2023 11:58:16	U	SMPL
PB151422BL - 1	PBW	-52	-0.0099	-	7470A	PPB	15 Mar 2023 12:00:33	U	SMPL
PB151422BS - 1	LCSW	18990	4.2677	-	7470A	PPB	15 Mar 2023 12:05:46	U	SMPL
O1863-04 - 1	030923-1	-2	0.0013	-	7470A	PPB	15 Mar 2023 12:08:02	U	SMPL
O1876-02 - 1	38414-CONCRETE	35	0.0096	-	7470A	PPB	15 Mar 2023 12:10:19	U	SMPL
O1877-03 - 1	COMP-SOIL	39	0.0105	-	7470A	PPB	15 Mar 2023 12:12:36	U	SMPL
O1898-02 - 1	SAMPLE-1-COMP	61	0.0155	-	7470A	PPB	15 Mar 2023 12:14:53	U	SMPL
O1900-04 - 1	HT-OILYDEBRIS-0313	98	0.0238	-	7470A	PPB	15 Mar 2023 12:17:08	U	SMPL
CCV67 - 1	CCV67	22093	4.9647	-	7470A	PPB	15 Mar 2023 12:19:24	U	SMPL
CCB67 - 1	CCB67	-35	-0.0061	-	7470A	PPB	15 Mar 2023 12:21:40	U	SMPL
O1900-07 - 1	HT-OIL-0313	965	0.2186	-	7470A	PPB	15 Mar 2023 12:23:55	U	SMPL
O1905-02 - 1	UNION-1	208	0.0485	-	7470A	PPB	15 Mar 2023 12:26:11	U	SMPL
O1905-05 - 1	MAPLE-1	172	0.0404	-	7470A	PPB	15 Mar 2023 12:28:27	U	SMPL
O1905-08 - 1	MAPLE-2	63	0.0159	-	7470A	PPB	15 Mar 2023 12:30:43	U	SMPL
O1905-11 - 1	MAPLE-3	84	0.0207	-	7470A	PPB	15 Mar 2023 12:32:58	U	SMPL
O1905-11DUP - 1	MAPLE-3DUP	125	0.0299	-	7470A	PPB	15 Mar 2023 12:35:15	U	SMPL
O1905-11MS - 1	MAPLE-3MS	17104	3.8440	-	7470A	PPB	15 Mar 2023 12:37:33	U	SMPL
O1905-11MSD - 1	MAPLE-3MSD	16733	3.7606	-	7470A	PPB	15 Mar 2023 12:39:50	U	SMPL
CCV68 - 1	CCV68	21609	4.8560	-	7470A	PPB	15 Mar 2023 12:42:08	U	SMPL
CCB68 - 1	CCB68	-61	-0.0119	-	7470A	PPB	15 Mar 2023 12:44:24	U	SMPL
PB151381TB - 1	PB151381TB	0	0.0018	-	7470A	PPB	15 Mar 2023 12:46:40	U	SMPL
O1905-11LX5 - 1		48	0.0126	-	7470A	PPB	15 Mar 2023 12:48:56	U	SMPL
O1905-11A - 1		18024	4.0507	-	7470A	PPB	15 Mar 2023 12:51:11	U	SMPL
CCV69 - 1	CCV69	21806	4.9002	-	7470A	PPB	15 Mar 2023 12:53:27	U	SMPL
CCB69 - 1	CCB69	4	0.0027	-	7470A	PPB	15 Mar 2023 12:55:42	U	SMPL