

LB124508


SFAM01.1

INSTRUMENT ID: CVA

Linear

μ Abs.:

50729



A= 0.0000e+000

B= 1.9744e-004 slope

C= 8.5284e-003 + intercept

Rho= 0.9999775

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.035	0.035	132	0.000	132					11
0.05	0.050					185	0				3
0.20	0.200	0.206	0.006	998	0.0 %	998					12
2.50	2.500	2.462	-0.038	12428	0.0 %	12428					0000
5.00	5.000	4.981	-0.019	25186	0.0 %	25186					
7.50	7.500	7.492	-0.008	37900	0.0 %	37900					
10.0	10.000	10.025	0.025	50729	0.0 %	50729					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	132	-		0 SFAM01.1	PPB	3/15/2023 14:05	Std
	0.2 S01	998	-		0.2 SFAM01.1	PPB	3/15/2023 14:07	Std
	2.5 S02	12428	-		2.5 SFAM01.1	PPB	3/15/2023 14:10	Std
	5 S03	25186	-		5 SFAM01.1	PPB	3/15/2023 14:12	Std
	7.5 S04	37900	-		7.5 SFAM01.1	PPB	3/15/2023 14:14	Std
	10 S05	50729	-		10 SFAM01.1	PPB	3/15/2023 14:19	Std
ICV087	ICV087	21224	4.199 -		SFAM01.1	PPB	3/15/2023 14:22	SMPL
ICB087	ICB087	5	0.0095 -		SFAM01.1	PPB	3/15/2023 14:24	SMPL
CCV070	CCV070	25116	4.9675 -		SFAM01.1	PPB	3/15/2023 14:26	SMPL
CCB070	CCB070	-9	0.0068 -		SFAM01.1	PPB	3/15/2023 14:29	SMPL
PB151448BL	PBS448	23	0.0131 -		SFAM01.1	PPB	3/15/2023 14:31	SMPL
O1851-01	MC0AC4	461	0.0995 -		SFAM01.1	PPB	3/15/2023 14:33	SMPL
O1851-02	MC0AC5	233	0.0545 -		SFAM01.1	PPB	3/15/2023 14:35	SMPL
O1851-03	MC0AC6	236	0.0551 -		SFAM01.1	PPB	3/15/2023 14:38	SMPL
O1851-04	MC0AC7	417	0.0909 -		SFAM01.1	PPB	3/15/2023 14:40	SMPL
O1851-05	MC0AC8	670	0.1408 -		SFAM01.1	PPB	3/15/2023 14:42	SMPL
O1851-06	MC0AC9	377	0.083 -		SFAM01.1	PPB	3/15/2023 14:44	SMPL
O1851-07	MC0AD0	440	0.0954 -		SFAM01.1	PPB	3/15/2023 14:47	SMPL
O1851-08	MC0AD1	-30	0.0026 -		SFAM01.1	PPB	3/15/2023 14:49	SMPL
O1851-09	MC0AD2	192	0.0464 -		SFAM01.1	PPB	3/15/2023 14:51	SMPL
O1851-10	MC0AD3	182	0.0445 -		SFAM01.1	PPB	3/15/2023 14:54	SMPL
O1851-11	MC0AD8	163	0.0407 -		SFAM01.1	PPB	3/15/2023 14:56	SMPL
O1851-12	MC0AD9	289	0.0656 -		SFAM01.1	PPB	3/15/2023 14:58	SMPL
O1851-13	MC0AE0	1234	0.2522 -		SFAM01.1	PPB	3/15/2023 15:00	SMPL
O1851-14	MC0AE0D	1242	0.2538 -		SFAM01.1	PPB	3/15/2023 15:05	SMPL
O1851-16	MC0AE1	608	0.1286 -		SFAM01.1	PPB	3/15/2023 15:10	SMPL
O1851-17	MC0AE2	513	0.1098 -		SFAM01.1	PPB	3/15/2023 15:12	SMPL
O1851-18	MC0AE3	1389	0.2828 -		SFAM01.1	PPB	3/15/2023 15:14	SMPL
O1851-19	MC0AE4	244	0.0567 -		SFAM01.1	PPB	3/15/2023 15:16	SMPL
CCV071	CCV071	25162	4.9766 -		SFAM01.1	PPB	3/15/2023 15:19	SMPL
CCB071	CCB071	-23	0.004 -		SFAM01.1	PPB	3/15/2023 15:21	SMPL
O1851-20	MC0AE5	298	0.0674 -		SFAM01.1	PPB	3/15/2023 15:23	SMPL
O1851-21	MC0AE6	406	0.0887 -		SFAM01.1	PPB	3/15/2023 15:26	SMPL
PB151449BL	PBS449	30	0.0145 -		SFAM01.1	PPB	3/15/2023 15:28	SMPL
O1888-01	YBQW7	998944	HIGH -		SFAM01.1	PPB	3/15/2023 15:30	
O1888-02	YBQW8	954307	HIGH -		SFAM01.1	PPB	3/15/2023 15:32	
O1888-06	YBQX0	99762	19.7058 -		SFAM01.1	PPB	3/15/2023 15:41	SMPL
O1888-07	YBQX2	710991	HIGH -		SFAM01.1	PPB	3/15/2023 15:43	
O1888-08	YBQX3	699954	HIGH -		SFAM01.1	PPB	3/15/2023 15:45	
O1888-09	YBQX4	311007	61.4145 -		SFAM01.1	PPB	3/15/2023 15:47	SMPL
O1888-10	YBQX5	277727	54.8436 -		SFAM01.1	PPB	3/15/2023 15:50	SMPL
O1888-11	YBQX6	191264	37.7721 -		SFAM01.1	PPB	3/15/2023 15:52	SMPL
O1888-12	YBQX7	211511	41.7698 -		SFAM01.1	PPB	3/15/2023 15:54	SMPL
O1888-13	YBQX1	100454	19.8424 -		SFAM01.1	PPB	3/15/2023 15:57	SMPL
O1888-14	YBQX8	20407	4.0377 -		SFAM01.1	PPB	3/15/2023 15:59	SMPL
O1888-15	YBQX9	175607	34.6808 -		SFAM01.1	PPB	3/15/2023 16:01	SMPL

LB124503
INSTRUMENT ID : CV1

O1888-16	YBQY0	105150	20.7696 -	SFAM01.1	PPB	3/15/2023 16:04	SMPL
O1888-17	YBQY5	146270	28.8884 -	SFAM01.1	PPB	3/15/2023 16:06	SMPL
CCV072	CCV072	23976	4.7424 -	SFAM01.1	PPB	3/15/2023 16:08	SMPL
CCB072	CCB072	-107	-0.0126 -	SFAM01.1	PPB	3/15/2023 16:11	SMPL
O1888-18	YBQY6	104853	20.711 -	SFAM01.1	PPB	3/15/2023 16:13	SMPL
O1888-19	YBQY8	134451	26.5549 -	SFAM01.1	PPB	3/15/2023 16:15	SMPL
O1888-20	YBQY9	103411	20.4262 -	SFAM01.1	PPB	3/15/2023 16:18	SMPL
O1888-21	YBQZ0	39567	7.8207 -	SFAM01.1	PPB	3/15/2023 16:20	SMPL
O1851-15	MCOAE0S	12817	2.5391 -	SFAM01.1	PPB	3/15/2023 16:22	SMPL
O1888-01DLX10		205349	40.5531 -	SFAM01.1	PPB	3/15/2023 16:24	SMPL
O1888-02DLX10		195229	38.555 -	SFAM01.1	PPB	3/15/2023 16:27	SMPL
O1888-03		154171	30.4484 -	SFAM01.1	PPB	3/15/2023 16:29	SMPL
O1888-04		152626	30.1434 -	SFAM01.1	PPB	3/15/2023 16:31	SMPL
O1888-05		171476	33.8652 -	SFAM01.1	PPB	3/15/2023 16:36	SMPL
O1888-03DLX5		31960	6.3188 -	SFAM01.1	PPB	3/15/2023 16:39	SMPL
O1888-04DLX5		33497	6.6223 -	SFAM01.1	PPB	3/15/2023 16:46	SMPL
O1888-05DLX5		29423	5.8179 -	SFAM01.1	PPB	3/15/2023 16:48	SMPL
O1888-06DLX5		19467	3.8521 -	SFAM01.1	PPB	3/15/2023 16:51	SMPL
O1888-07DLX10		119191	23.5419 -	SFAM01.1	PPB	3/15/2023 16:53	SMPL
O1888-08DLX10		121438	23.9855 -	SFAM01.1	PPB	3/15/2023 16:55	SMPL
O1888-09DLX10		33369	6.597 -	SFAM01.1	PPB	3/15/2023 16:57	SMPL
CCV073	CCV073	24723	4.8899 -	SFAM01.1	PPB	3/15/2023 17:00	SMPL
CCB073	CCB073	-104	-0.012 -	SFAM01.1	PPB	3/15/2023 17:02	SMPL
O1888-10DLX10		26731	5.2864 -	SFAM01.1	PPB	3/15/2023 17:04	SMPL
O1888-12DLX10		20931	4.1412 -	SFAM01.1	PPB	3/15/2023 17:09	SMPL
O1888-13DLX5		20021	3.9615 -	SFAM01.1	PPB	3/15/2023 17:11	SMPL
O1888-15DLX10		17844	3.5317 -	SFAM01.1	PPB	3/15/2023 17:13	SMPL
O1888-16DLX5		21401	4.234 -	SFAM01.1	PPB	3/15/2023 17:16	SMPL
O1888-17DLX5		29462	5.8256 -	SFAM01.1	PPB	3/15/2023 17:18	SMPL
O1888-18DLX5		20042	3.9657 -	SFAM01.1	PPB	3/15/2023 17:20	SMPL
O1888-01DL2X100		20549	4.0658 -	SFAM01.1	PPB	3/15/2023 17:27	SMPL
O1888-02DL2X100		17817	3.5264 -	SFAM01.1	PPB	3/15/2023 17:29	SMPL
O1888-07DL2X50		23492	4.6468 -	SFAM01.1	PPB	3/15/2023 17:31	SMPL
O1888-08DL2X50		24386	4.8234 -	SFAM01.1	PPB	3/15/2023 17:34	SMPL
O1888-19DLX5		26807	5.3014 -	SFAM01.1	PPB	3/15/2023 17:38	SMPL
O1888-20DLX5		21451	4.2439 -	SFAM01.1	PPB	3/15/2023 17:40	SMPL
O1888-11DLX10		19654	3.8891 -	SFAM01.1	PPB	3/15/2023 17:43	SMPL
CCV074	CCV074	24648	4.8751 -	SFAM01.1	PPB	3/15/2023 17:45	SMPL
CCB074	CCB074	-9	0.0068 -	SFAM01.1	PPB	3/15/2023 17:47	SMPL