

LB124518


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49481



A= 0.0000e+000

B= 2.0449e-004 *Slope*C= -7.7588e-002 *y-intercept*

Rho= 0.9998689

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>σ/μ</i>
0.00	0.000	0.018	0.018	468	0.000	468					1
0.05	0.050					185	0				6
0.20	0.200	0.212	0.012	1414	0.0 %	1414					2
2.50	2.500	2.439	-0.061	12306	0.0 %	12306					2
5.00	5.000	5.083	0.083	25239	0.0 %	25239					1
7.50	7.500	7.407	-0.093	36603	0.0 %	36603					0
10.0	10.000	10.041	0.041	49481	0.0 %	49481					

LB124518

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	468	-		0 SFAM01.1	PPB	3/16/2023 12:27	Std
	0.2 S01	1414	-		0.2 SFAM01.1	PPB	3/16/2023 12:30	Std
	2.5 S02	12306	-		2.5 SFAM01.1	PPB	3/16/2023 12:32	Std
	5 S03	25239	-		5 SFAM01.1	PPB	3/16/2023 12:34	Std
	7.5 S04	36603	-		7.5 SFAM01.1	PPB	3/16/2023 12:36	Std
	10 S05	49481	-		10 SFAM01.1	PPB	3/16/2023 12:41	Std
ICV088	ICV088	18891	3.7854 -		SFAM01.1	PPB	3/16/2023 12:47	SMPL
ICB088	ICB088	384	0.0009 -		SFAM01.1	PPB	3/16/2023 12:49	SMPL
CCV075	CCV075	24961	5.0266 -		SFAM01.1	PPB	3/16/2023 12:51	SMPL
CCB075	CCB075	289	-0.0185 -		SFAM01.1	PPB	3/16/2023 12:54	SMPL
PB151475BL	PBS475	58	-0.0657 -		SFAM01.1	PPB	3/16/2023 12:56	SMPL
O1889-01	YBQY1	419036	HIGH -		SFAM01.1	PPB	3/16/2023 12:58	
O1889-02	YBQY2	154424	31.5003 -		SFAM01.1	PPB	3/16/2023 13:00	SMPL
O1889-03	YBQY3	315372	64.4123 -		SFAM01.1	PPB	3/16/2023 13:02	SMPL
O1889-04	YBQY4	143655	29.2982 -		SFAM01.1	PPB	3/16/2023 13:04	SMPL
O1889-05	YBQY4D	149270	30.4464 -		SFAM01.1	PPB	3/16/2023 13:07	SMPL
O1889-06	YBQY4S	152446	31.0958 -		SFAM01.1	PPB	3/16/2023 13:12	SMPL
O1889-07	YBQY7	62756	12.7553 -		SFAM01.1	PPB	3/16/2023 13:14	SMPL
O1889-08	YBQZ1	116055	23.6543 -		SFAM01.1	PPB	3/16/2023 13:17	SMPL
O1889-09	YBQZ2	147092	30.001 -		SFAM01.1	PPB	3/16/2023 13:19	SMPL
O1889-10	YBQZ3	85914	17.4908 -		SFAM01.1	PPB	3/16/2023 13:21	SMPL
O1889-11	YBQZ4	5338	1.014 -		SFAM01.1	PPB	3/16/2023 13:24	SMPL
O1889-12	YBQZ7	72570	14.7621 -		SFAM01.1	PPB	3/16/2023 13:26	SMPL
O1911-01	YBQZ5	295237	HIGH -		SFAM01.1	PPB	3/16/2023 13:30	
O1911-02	YBQZ6	121096	24.6851 -		SFAM01.1	PPB	3/16/2023 13:32	SMPL
O1911-03	YBQZ8	212743	43.4259 -		SFAM01.1	PPB	3/16/2023 13:35	SMPL
O1911-04	YBQZ9	266866	54.4934 -		SFAM01.1	PPB	3/16/2023 13:37	SMPL
O1911-05	YBR00	234063	47.7856 -		SFAM01.1	PPB	3/16/2023 13:39	SMPL
O1911-06	YBR01	165010	33.665 -		SFAM01.1	PPB	3/16/2023 13:42	SMPL
CCV076	CCV076	24563	4.9453 -		SFAM01.1	PPB	3/16/2023 13:44	SMPL
CCB076	CCB076	227	-0.0312 -		SFAM01.1	PPB	3/16/2023 13:46	SMPL
O1911-07	YBR02	240887	49.181 -		SFAM01.1	PPB	3/16/2023 13:49	SMPL
O1911-08	YBR03	455323	HIGH -		SFAM01.1	PPB	3/16/2023 13:51	
O1911-09	YBR03D	447431	HIGH -		SFAM01.1	PPB	3/16/2023 13:53	
O1911-10	YBR03S	471788	HIGH -		SFAM01.1	PPB	3/16/2023 13:55	
O1911-11	YBR04	138634	28.2714 -		SFAM01.1	PPB	3/16/2023 13:56	SMPL
O1911-12	YBR05	243769	49.7703 -		SFAM01.1	PPB	3/16/2023 13:59	SMPL
O1911-13	YBR66	435503	HIGH -		SFAM01.1	PPB	3/16/2023 14:01	
O1911-14	YBR67	283007	57.794 -		SFAM01.1	PPB	3/16/2023 14:03	SMPL
O1911-15	YBR68	297335	HIGH -		SFAM01.1	PPB	3/16/2023 14:05	
O1911-16	YBR69	1232033	HIGH -		SFAM01.1	PPB	3/16/2023 14:07	
O1911-17	YBR70	488813	HIGH -		SFAM01.1	PPB	3/16/2023 14:09	
O1911-18	YBR71	326927	HIGH -		SFAM01.1	PPB	3/16/2023 14:11	
O1911-19	YBR72	271663	55.4743 -		SFAM01.1	PPB	3/16/2023 14:13	SMPL
O1911-20	YBR73	540948	HIGH -		SFAM01.1	PPB	3/16/2023 14:16	
O1911-21	YBR74	557766	HIGH -		SFAM01.1	PPB	3/16/2023 14:18	

LB124518
INSTRUMENT ID : CV1

O1911-22	YBR75	371343	HIGH	-	SFAM01.1	PPB	3/16/2023 14:20
PB151477BL	PBS477	463	0.0171	-	SFAM01.1	PPB	3/16/2023 14:24 SMPL
O1913-01	YBR06	296291	HIGH	-	SFAM01.1	PPB	3/16/2023 14:26
O1913-02	YBR07	560937	HIGH	-	SFAM01.1	PPB	3/16/2023 14:28
O1913-03	YBR08	497423	HIGH	-	SFAM01.1	PPB	3/16/2023 14:30
CCV077	CCV077	23703	4.7694	-	SFAM01.1	PPB	3/16/2023 14:32 SMPL
CCB077	CCB077	183	-0.0402	-	SFAM01.1	PPB	3/16/2023 14:34 SMPL
O1913-04	YBR09	131266	26.7648	-	SFAM01.1	PPB	3/16/2023 14:37 SMPL
O1913-05	YBR10	310446	63.405	-	SFAM01.1	PPB	3/16/2023 14:39 SMPL
O1913-06	YBR11	291593	59.5498	-	SFAM01.1	PPB	3/16/2023 14:41 SMPL
O1913-07	YBR11D	289240	HIGH	-	SFAM01.1	PPB	3/16/2023 14:44
O1913-08	YBR11S	300150	HIGH	-	SFAM01.1	PPB	3/16/2023 14:46
O1913-09	YBR12	170779	34.8447	-	SFAM01.1	PPB	3/16/2023 14:48 SMPL
O1913-10	YBR13	306470	HIGH	-	SFAM01.1	PPB	3/16/2023 14:50
O1913-11	YBR14	121902	24.8499	-	SFAM01.1	PPB	3/16/2023 14:52 SMPL
O1913-12	YBR65	657750	HIGH	-	SFAM01.1	PPB	3/16/2023 14:54
O1913-13	YBR76	252610	51.5782	-	SFAM01.1	PPB	3/16/2023 14:56 SMPL
O1889-01DLX10		55133	11.1965	-	SFAM01.1	PPB	3/16/2023 14:58 SMPL
O1889-03DLX10		18331	3.6709	-	SFAM01.1	PPB	3/16/2023 15:03 SMPL
O1889-04DLX5		28096	5.6677	-	SFAM01.1	PPB	3/16/2023 15:06 SMPL
O1889-05DLX5		29124	5.8779	-	SFAM01.1	PPB	3/16/2023 15:08 SMPL
O1889-06DLX5		31838	6.4329	-	SFAM01.1	PPB	3/16/2023 15:10 SMPL
O1889-07DLX2		33117	6.6945	-	SFAM01.1	PPB	3/16/2023 15:13 SMPL
O1889-08DLX5		22152	4.4522	-	SFAM01.1	PPB	3/16/2023 15:15 SMPL
O1889-09DLX10		15616	3.1157	-	SFAM01.1	PPB	3/16/2023 15:17 SMPL
O1889-10DLX5		18527	3.711	-	SFAM01.1	PPB	3/16/2023 15:19 SMPL
CCV078	CCV078	24728	4.979	-	SFAM01.1	PPB	3/16/2023 15:22 SMPL
CCB078	CCB078	307	-0.0148	-	SFAM01.1	PPB	3/16/2023 15:24 SMPL
O1889-12DLX5		14234	2.8331	-	SFAM01.1	PPB	3/16/2023 15:26 SMPL
PB151476BL	PBW476	332	-0.0097	-	SFAM01.1	PPB	3/16/2023 15:29 SMPL
O1911-01DLX10		38902	7.8774	-	SFAM01.1	PPB	3/16/2023 15:31 SMPL
O1911-02DLX5		23955	4.8209	-	SFAM01.1	PPB	3/16/2023 15:33 SMPL
O1911-03DLX10		22244	4.471	-	SFAM01.1	PPB	3/16/2023 15:35 SMPL
O1911-04DLX10		28567	5.764	-	SFAM01.1	PPB	3/16/2023 15:38 SMPL
O1911-05DLX10		24261	4.8835	-	SFAM01.1	PPB	3/16/2023 15:40 SMPL
O1911-06DLX10		16821	3.3621	-	SFAM01.1	PPB	3/16/2023 15:42 SMPL
O1911-07DLX10		25242	5.0841	-	SFAM01.1	PPB	3/16/2023 15:45 SMPL
O1911-08DLX10		71480	14.5392	-	SFAM01.1	PPB	3/16/2023 15:47 SMPL
O1911-09DLX10		66206	13.4608	-	SFAM01.1	PPB	3/16/2023 15:49 SMPL
O1911-10DLX10		73451	14.9423	-	SFAM01.1	PPB	3/16/2023 15:51 SMPL
O1911-11DLX5		26073	5.254	-	SFAM01.1	PPB	3/16/2023 15:54 SMPL
O1911-12DLX10		25313	5.0986	-	SFAM01.1	PPB	3/16/2023 15:56 SMPL
O1911-13DLX10		62580	12.7193	-	SFAM01.1	PPB	3/16/2023 15:58 SMPL
O1911-14DLX10		31301	6.3231	-	SFAM01.1	PPB	3/16/2023 16:01 SMPL
O1911-15DLX10		40390	8.1817	-	SFAM01.1	PPB	3/16/2023 16:03 SMPL
O1911-16DLX10		326012	66.5881	-	SFAM01.1	PPB	3/16/2023 16:05 SMPL
O1911-17DLX10		74203	15.0961	-	SFAM01.1	PPB	3/16/2023 16:08 SMPL

LB124518
INSTRUMENT ID : CV1

O1911-18DLX10		45389	9.2039 -	SFAM01.1	PPB	3/16/2023 16:10 SMPL
CCV079	CCV079	24668	4.9667 -	SFAM01.1	PPB	3/16/2023 16:12 SMPL
CCB079	CCB079	279	-0.0205 -	SFAM01.1	PPB	3/16/2023 16:15 SMPL
O1911-19DLX10		26326	5.3058 -	SFAM01.1	PPB	3/16/2023 16:17 SMPL
O1911-20DLX10		75405	15.3419 -	SFAM01.1	PPB	3/16/2023 16:19 SMPL
O1911-21DLX10		75115	15.2826 -	SFAM01.1	PPB	3/16/2023 16:21 SMPL
O1911-22DLX10		48739	9.889 -	SFAM01.1	PPB	3/16/2023 16:24 SMPL
O1913-01DLX10		37520	7.5948 -	SFAM01.1	PPB	3/16/2023 16:26 SMPL
O1913-02DLX10		93665	19.0758 -	SFAM01.1	PPB	3/16/2023 16:28 SMPL
O1913-03DLX10		76724	15.6116 -	SFAM01.1	PPB	3/16/2023 16:31 SMPL
O1913-04DLX5		27326	5.5103 -	SFAM01.1	PPB	3/16/2023 16:33 SMPL
O1913-05DLX10		34251	6.9263 -	SFAM01.1	PPB	3/16/2023 16:35 SMPL
O1913-06DLX10		32290	6.5253 -	SFAM01.1	PPB	3/16/2023 16:38 SMPL
O1913-07DLX10		39749	8.0506 -	SFAM01.1	PPB	3/16/2023 16:40 SMPL
O1913-08DLX10		44234	8.9677 -	SFAM01.1	PPB	3/16/2023 16:42 SMPL
O1913-09DLX10		17392	3.4789 -	SFAM01.1	PPB	3/16/2023 16:44 SMPL
O1913-10DLX10		44530	9.0283 -	SFAM01.1	PPB	3/16/2023 16:47 SMPL
O1913-11DLX5		25330	5.1021 -	SFAM01.1	PPB	3/16/2023 16:49 SMPL
O1913-12DLX10		102870	20.9581 -	SFAM01.1	PPB	3/16/2023 16:51 SMPL
O1913-13DLX10		19402	3.8899 -	SFAM01.1	PPB	3/16/2023 16:54 SMPL
CCV080	CCV080	24374	4.9066 -	SFAM01.1	PPB	3/16/2023 16:56 SMPL
CCB080	CCB080	187	-0.0393 -	SFAM01.1	PPB	3/16/2023 16:58 SMPL