



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124525

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	03/16/2023	14:10
2	CAL2	1	Water	NA	NA	20.6	7.00	03/16/2023	14:11
3	CAL3	1	Water	NA	NA	20.6	10.03	03/16/2023	14:15
4	ICV	1	Water	NA	NA	20.7	6.99	03/16/2023	14:20
5	CCV1	1	Water	NA	NA	20.7	2.01	03/16/2023	14:22
6	O1935-02	1	Water	NA	NA	20.4	11.32	03/16/2023	14:25
7	O1946-02	1	Water	NA	NA	20.1	9.45	03/16/2023	14:30
8	O1946-03	1	Water	NA	NA	20.2	6.06	03/16/2023	14:35
9	O1946-03DUP	1	Water	NA	NA	20.1	6.47	03/16/2023	14:36
10	CCV2	1	Water	NA	NA	20.6	12.02	03/16/2023	14:40

124525

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1946 W

WorkList ID : 168213

Department : Wet-Chemistry

Date : 03-16-2023 15:30:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/15/2023	Water	O1935-02	pH	Cool 4 deg C	PSEG03	J11	2673-75	03/15/2023	9040C
03/16/2023	Water	O1946-02	pH	Cool 4 deg C	PSEG03	J11	NWB-1895	03/16/2023	9040C
03/16/2023	Water	O1946-03	pH	Cool 4 deg C	PSEG03	J11	NWB-1896	03/16/2023	9040C

Date/Time 03-16-23 14:00
Raw Sample Received by: JPC (sm)
Raw Sample Relinquished by: JPC (sm)

Date/Time 03-16-23 16:30
Raw Sample Received by: JPC (sm)
Raw Sample Relinquished by: JPC (sm)