



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124527

Slope : 99.1

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	03/16/2023	14:10
2	CAL2	1	Water	NA	NA	20.6	7.00	03/16/2023	14:11
3	CAL3	1	Water	NA	NA	20.6	10.03	03/16/2023	14:15
4	ICV	1	Water	NA	NA	20.7	6.99	03/16/2023	14:20
5	CCV1	1	Water	NA	NA	20.7	2.02	03/16/2023	15:00
6	O1939-06	1	Water	NA	NA	20.7	8.06	03/16/2023	15:05
7	O1942-02	1	Water	NA	NA	20.7	6.95	03/16/2023	15:10
8	O1942-02DUP	1	Water	NA	NA	20.8	6.96	03/16/2023	15:11
9	CCV2	1	Water	NA	NA	20.7	12.03	03/16/2023	15:15

WORKLIST(Hardcopy Internal Chain)

UP 124527

WorkList Name : CORROSIONITY O1942 W WorkList ID : 168205 Department : Wet-Chemistry Date : 03-16-2023 10:36:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/15/2023	Water	O1939-06	Corrosivity	Cool 4 deg C	PSEG03	J11	D2958	03/15/2023	9040C
03/15/2023	Water	O1942-02	Corrosivity	Cool 4 deg C	PSEG03	J11	349	03/15/2023	9040C

Date/Time 03-16-23 10:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 03-16-23 16:30
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC