



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124531

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/16/2023	11:30
2	O1907-08		1	12	100.00	>212.0	>212.0	03/16/2023	12:00
3	O1907-08DUP		1	12	100.00	>212.0	>212.0	03/16/2023	12:30
4	O1923-01		1	14	100.00	>212.0	>212.0	03/16/2023	13:00
5	O1939-06		1	14	100.00	>212.0	>212.0	03/16/2023	13:30
6	O1942-02		1	14	100.00	>212.0	>212.0	03/16/2023	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-3-15

WorkList ID : 168153

Department : Wet-Chemistry

Date : 03-15-2023 08:23:44

66124531

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2023	Water	O1907-08	Flash Point	Cool 4 deg C	PSEG03	J11	33969	03/14/2023	1010B
03/20/2023	Water	O1939-06	Flash Point	Cool 4 deg C	PSEG03	J11	D2958	03/15/2023	1010B
03/20/2023	Water	O1942-02	Flash Point	Cool 4 deg C	PSEG03	J11	349	03/15/2023	1010B
03/15/2023	Water	O1923-01	Flash Point	Cool 4 deg C	SOUN03	L31	SAMPLE-1	03/15/2023	1010B

Date/Time 03/16/2023 11:00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JDC

Date/Time 03/16/2023 14:40
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: RM CWC