


LB124535

SFAM01.1

INSTRUMENT ID: 101

Linear



A= 0.0000e+000

B= 2.0194e-004 slope

C= -4.9137e-002 y-intercept

Rho= 0.9998468

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.044	0.044	459	0.000	459					0.00
0.05	0.050					185	0				0.00
0.20	0.200	0.212	0.012	1291	0.0 %	1291					0.00
2.50	2.500	2.421	-0.079	12234	0.0 %	12234					0.00
5.00	5.000	5.062	0.062	25310	0.0 %	25310					0.00
7.50	7.500	7.403	-0.097	36901	0.0 %	36901					0.00
10.0	10.000	10.059	0.059	50056	0.0 %	50056					0.00

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	459	-		0 SFAM01.1	PPB	3/17/2023 7:52	Std
	0.2 S01	1291	-		0.2 SFAM01.1	PPB	3/17/2023 7:54	Std
	2.5 S02	12234	-		2.5 SFAM01.1	PPB	3/17/2023 7:57	Std
	5 S03	25310	-		5 SFAM01.1	PPB	3/17/2023 7:59	Std
	7.5 S04	36901	-		7.5 SFAM01.1	PPB	3/17/2023 8:01	Std
	10 S05	50056	-		10 SFAM01.1	PPB	3/17/2023 8:04	Std
ICV089	ICV089	19631	3.9151 -		SFAM01.1	PPB	3/17/2023 8:09	SMPL
ICB089	ICB089	304	0.0123 -		SFAM01.1	PPB	3/17/2023 8:11	SMPL
CCV081	CCV081	25476	5.0954 -		SFAM01.1	PPB	3/17/2023 8:13	SMPL
CCB081	CCB081	348	0.0211 -		SFAM01.1	PPB	3/17/2023 8:16	SMPL
O1889-01DL2X25	YBQY1	24951	4.9894 -		SFAM01.1	PPB	3/17/2023 8:18	SMPL
O1889-02DLX10	YBQY2	17664	3.5179 -		SFAM01.1	PPB	3/17/2023 8:20	SMPL
O1911-01DL2X25	YBQZ5	18469	3.6804 -		SFAM01.1	PPB	3/17/2023 8:22	SMPL
O1911-08DL2X25	YBR03	33809	6.7782 -		SFAM01.1	PPB	3/17/2023 8:25	SMPL
O1911-09DL2X25	YBR03D	30150	6.0393 -		SFAM01.1	PPB	3/17/2023 8:27	SMPL
O1911-10DL2X25	YBR03S	32722	6.5587 -		SFAM01.1	PPB	3/17/2023 8:29	SMPL
O1911-13DL2X25	YBR66	28495	5.7051 -		SFAM01.1	PPB	3/17/2023 8:35	SMPL
O1911-15DL2X25	YBR68	15167	3.0137 -		SFAM01.1	PPB	3/17/2023 8:37	SMPL
O1911-16DL2X100	YBR69	35385	7.0964 -		SFAM01.1	PPB	3/17/2023 8:39	SMPL
O1911-18DL2X25	YBR71	18645	3.716 -		SFAM01.1	PPB	3/17/2023 8:46	SMPL
CCV082	CCV082	25858	5.1726 -		SFAM01.1	PPB	3/17/2023 8:49	SMPL
CCB082	CCB082	346	0.0207 -		SFAM01.1	PPB	3/17/2023 8:51	SMPL
O1911-22DL2X25	YBR75	23420	4.6802 -		SFAM01.1	PPB	3/17/2023 8:58	SMPL
O1913-01DL2X25	YBR06	18332	3.6528 -		SFAM01.1	PPB	3/17/2023 9:00	SMPL
O1913-02DL2X50	YBR07	19121	3.8121 -		SFAM01.1	PPB	3/17/2023 9:05	SMPL
O1913-03DL2X25	YBR08	31642	6.3406 -		SFAM01.1	PPB	3/17/2023 9:08	SMPL
O1913-06DL2X25	YBR11	15334	3.0474 -		SFAM01.1	PPB	3/17/2023 9:11	SMPL
O1913-07DL2X25	YBR11D	14151	2.8085 -		SFAM01.1	PPB	3/17/2023 9:14	SMPL
O1913-08DL2X25	YBR11S	12965	2.569 -		SFAM01.1	PPB	3/17/2023 9:16	SMPL
O1913-10DL2X25	YBR13	18995	3.7867 -		SFAM01.1	PPB	3/17/2023 9:18	SMPL
O1913-12DL2X50	YBR65	21010	4.1936 -		SFAM01.1	PPB	3/17/2023 9:21	SMPL
O1913-13DLX10	YBR76	29834	5.9755 -		SFAM01.1	PPB	3/17/2023 9:23	SMPL
O1911-17DL2X25	YBR70	30774	6.1653 -		SFAM01.1	PPB	3/17/2023 9:25	SMPL
O1911-20DL2X25	YBR73	28863	5.7794 -		SFAM01.1	PPB	3/17/2023 9:27	SMPL
O1911-21DL2X25	YBR74	37262	7.4755 -		SFAM01.1	PPB	3/17/2023 9:30	SMPL
CCV083	CCV083	25186	5.0369 -		SFAM01.1	PPB	3/17/2023 9:32	SMPL
CCB083	CCB083	327	0.0169 -		SFAM01.1	PPB	3/17/2023 9:34	SMPL