


LB1241-39

245.1

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 2.2053e-004

C= -9.2990e-003

Rho= 0.9997638

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	0.014	0.014	107	0.000	107					0.00
0.05	0.050	0.042	-0.008	231	0.0 %	231					-16
0.2	0.200	0.171	-0.029	817	0.0 %	817	0				-15
2.5	2.500	2.545	0.045	11582	0.0 %	11582					2
5.0	5.000	5.074	0.074	23052	0.0 %	23052					1
7.5	7.500	7.325	-0.175	33257	0.0 %	33257					-2
10.0	10.000	10.079	0.079	45745	0.0 %	45745					1

LB124539 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Mar 2023 11:05:16

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.0000	245.1	PPB	17 Mar 2023 11:07:35	S	Std
0.05 - 1	50.05	231	-	0.0500	245.1	PPB	17 Mar 2023 11:09:53	S	Std
0.2 - 1	50.2	817	-	0.2000	245.1	PPB	17 Mar 2023 11:12:10	S	Std
2.5 - 1	52.5	11582	-	2.5000	245.1	PPB	17 Mar 2023 11:14:27	S	Std
5.0 - 1	55	23052	-	5.0000	245.1	PPB	17 Mar 2023 11:19:20	S	Std
7.5 - 1	57.5	33257	-	7.5000	245.1	PPB	17 Mar 2023 11:21:36	S	Std
10.0 - 1	510	45745	-	10.0000	245.1	PPB	17 Mar 2023 11:26:28	S	Std
ICV90 - 1	ICV90	18596	4.0917	-245.1		PPB	17 Mar 2023 11:29:09	U	SMPL
ICB90 - 1	ICB90	-14	-0.0124	-245.1		PPB	17 Mar 2023 11:31:24	U	SMPL
CCV84 - 1	CCV84	23083	5.0812	-245.1		PPB	17 Mar 2023 11:35:36	U	SMPL
CCB84 - 1	CCB84	-28	-0.0155	-245.1		PPB	17 Mar 2023 11:37:51	U	SMPL
CRA - 1	CRA	239	0.0434	-245.1		PPB	17 Mar 2023 11:45:15	U	SMPL
HighStd - 1	HighStd	45665	10.0613	-245.1		PPB	17 Mar 2023 11:47:30	U	SMPL
ChkStd - 1	ChkStd	32065	7.0665	-245.1		PPB	17 Mar 2023 11:49:46	U	SMPL
PB151473BL - 1	PBW	139	0.0214	-245.1		PPB	17 Mar 2023 11:52:03	U	SMPL
PB151473BS - 1	LCSW	19308	4.2487	-245.1		PPB	17 Mar 2023 11:54:21	U	SMPL
O1627-12 - 1	MDL-WATER-06-QT1-2023	219	0.0390	-245.1		PPB	17 Mar 2023 11:58:55	U	SMPL
O1923-01 - 1	SAMPLE-1	121	0.0174	-245.1		PPB	17 Mar 2023 12:01:45	U	SMPL
O1923-01DUP - 1	SAMPLE-1DUP	119	0.0169	-245.1		PPB	17 Mar 2023 12:04:02	U	SMPL
O1923-01MS - 1	SAMPLE-1MS	12148	2.6697	-245.1		PPB	17 Mar 2023 12:06:18	U	SMPL
CCV85 - 1	CCV85	23013	5.0658	-245.1		PPB	17 Mar 2023 12:08:33	U	SMPL
CCB85 - 1	CCB85	-31	-0.0161	-245.1		PPB	17 Mar 2023 12:10:50	U	SMPL
O1923-01MSD - 1	SAMPLE-1MSD	14212	3.1249	-245.1		PPB	17 Mar 2023 12:13:06	U	SMPL
O1627-10 - 1	LOD-MDL-WATER-04-QT1-2023	226	0.0405	-245.1		PPB	17 Mar 2023 12:18:27	U	SMPL
O1627-10RE - 1	LOD-MDL-WATER-04-QT1-2023RE	793	0.1656	-245.1		PPB	17 Mar 2023 12:20:42	U	SMPL
O1923-01LX5 - 1		1	-0.0091	-245.1		PPB	17 Mar 2023 12:22:58	U	SMPL
O1923-01A - 1		18878	4.1539	-245.1		PPB	17 Mar 2023 12:25:15	U	SMPL
CCV86 - 1	CCV86	22549	4.9635	-245.1		PPB	17 Mar 2023 12:27:31	U	SMPL
CCB86 - 1	CCB86	-35	-0.0170	-245.1		PPB	17 Mar 2023 12:29:47	U	SMPL