



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB124541

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/17/2023	11:10
2	CAL2	1	Water	NA	NA	20.7	7.00	03/17/2023	11:11
3	CAL3	1	Water	NA	NA	20.7	10.03	03/17/2023	11:12
4	ICV	1	Water	NA	NA	20.7	6.99	03/17/2023	11:15
5	CCV1	1	Water	NA	NA	20.7	2.01	03/17/2023	11:20
6	O1967-01	1	Water	NA	NA	23.1	10.23	03/17/2023	11:25
7	O1967-01DUP	1	Water	NA	NA	23.2	10.24	03/17/2023	11:26
8	O1967-02	1	Water	NA	NA	23.1	10.00	03/17/2023	11:30
9	O1967-03	1	Water	NA	NA	23.3	10.31	03/17/2023	11:33
10	O1967-04	1	Water	NA	NA	22.8	10.31	03/17/2023	11:37
11	O1967-05	1	Water	NA	NA	22.6	8.40	03/17/2023	11:40
12	O1967-06	1	Water	NA	NA	22.6	7.73	03/17/2023	11:45
13	O1967-07	1	Water	NA	NA	23.3	7.51	03/17/2023	11:48
14	O1967-08	1	Water	NA	NA	22.8	7.34	03/17/2023	11:50
15	O1967-09	1	Water	NA	NA	22.4	7.42	03/17/2023	12:00
16	CCV2	1	Water	NA	NA	20.7	12.01	03/17/2023	12:03
17	O1967-10	1	Water	NA	NA	23.1	7.29	03/17/2023	12:05
18	CCV3	1	Water	NA	NA	20.6	2.02	03/17/2023	12:10

WORKLIST(Hardcopy Internal Chain)

12541

WorkList Name : ph o1967

WorkList ID : 168241

Department : Wet-Chemistry

Date : 03-17-2023 10:52:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/16/2023	Water	O1967-01	pH	✓	Cool 4 deg C	ATCE02 J11	3X1	03/16/2023	9040C
03/16/2023	Water	O1967-02	pH	✓	Cool 4 deg C	ATCE02 J11	3X2	03/16/2023	9040C
03/16/2023	Water	O1967-03	pH	✓	Cool 4 deg C	ATCE02 J11	3X3	03/16/2023	9040C
03/16/2023	Water	O1967-04	pH	✓	Cool 4 deg C	ATCE02 J11	3X4	03/16/2023	9040C
03/16/2023	Water	O1967-05	pH	✓	Cool 4 deg C	ATCE02 J11	3X5	03/16/2023	9040C
03/16/2023	Water	O1967-06	pH	✓	Cool 4 deg C	ATCE02 J11	3X6	03/16/2023	9040C
03/16/2023	Water	O1967-07	pH	✓	Cool 4 deg C	ATCE02 J11	3X7	03/16/2023	9040C
03/16/2023	Water	O1967-08	pH	✓	Cool 4 deg C	ATCE02 J11	3X8	03/16/2023	9040C
03/16/2023	Water	O1967-09	pH	✓	Cool 4 deg C	ATCE02 J11	3X9	03/16/2023	9040C
03/16/2023	Water	O1967-10	pH	✓	Cool 4 deg C	ATCE02 J11	3X10	03/16/2023	9040C

Date/Time 03-17-23 11:00

Raw Sample Received by: 10100C

Raw Sample Relinquished by: RM CWC

Date/Time 03-17-23 13:00

Raw Sample Received by: RM CWC

Raw Sample Relinquished by: 10100C