


68124543

7470A

INSTRUMENT ID: CV2

μ Abs.:

45989



Linear

A= 0.0000e+000

B= 2.1900e-004

C= -1.5458e-002

Rho= 0.9996127

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.027	0.027	192	0.000	192					
0.2	0.200	0.142	-0.058	721	0.0 %	721					-29
2.5	2.500	2.667	0.167	12251	0.0 %	12251					7
5.0	5.000	4.835	-0.165	22147	0.0 %	22147					-3
7.5	7.500	7.473	-0.027	34194	0.0 %	34194					0
10.0	10.000	10.056	0.056	45989	0.0 %	45989					1

LB124543 INSTRUMENT IOD : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Mar 2023 12:45:40

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	192	-	0.0000	7470A	PPB	17 Mar 2023 12:47:27	S	Std
0.2 - 1	SD-2	721	-	0.2000	7470A	PPB	17 Mar 2023 12:49:45	S	Std
2.5 - 1	SD-5	12251	-	2.5000	7470A	PPB	17 Mar 2023 12:52:01	S	Std
5.0 - 1	SD	22147	-	5.0000	7470A	PPB	17 Mar 2023 12:54:19	S	Std
7.5 - 1	SD-5	34194	-	7.5000	7470A	PPB	17 Mar 2023 12:56:35	S	Std
10.0 - 1	SD	45989	-	10.0000	7470A	PPB	17 Mar 2023 13:04:04	S	Std
ICV91 - 1	ICV91	18920	4.1280	-	7470A	PPB	17 Mar 2023 13:06:43	U	SMPL
ICB91 - 1	ICB91	36	-0.0076	-	7470A	PPB	17 Mar 2023 13:08:59	U	SMPL
CCV87 - 1	CCV87	21847	4.7690	-	7470A	PPB	17 Mar 2023 13:11:14	U	SMPL
CCB87 - 1	CCB87	36	-0.0076	-	7470A	PPB	17 Mar 2023 13:13:30	U	SMPL
CRA - 1	CRA	901	0.1819	-	7470A	PPB	17 Mar 2023 13:15:46	U	SMPL
HighStd - 1	HighStd	45899	10.0362	-	7470A	PPB	17 Mar 2023 13:18:01	U	SMPL
ChkStd - 1	ChkStd	32100	7.0143	-	7470A	PPB	17 Mar 2023 13:20:17	U	SMPL
PB151441BL - 1	PBW	-72	-0.0312	-	7470A	PPB	17 Mar 2023 13:22:34	U	SMPL
PB151441BS - 1	LCSW	14995	3.2684	-	7470A	PPB	17 Mar 2023 13:24:51	U	SMPL
O1907-01 - 1	33956	35	-0.0078	-	7470A	PPB	17 Mar 2023 13:27:08	U	SMPL
O1907-02 - 1	33957	90	0.0043	-	7470A	PPB	17 Mar 2023 13:29:24	U	SMPL
O1907-03 - 1	33958	104	0.0073	-	7470A	PPB	17 Mar 2023 13:31:41	U	SMPL
O1907-04 - 1	33959	231	0.0351	-	7470A	PPB	17 Mar 2023 13:33:59	U	SMPL
O1907-05 - 1	33960	187	0.0255	-	7470A	PPB	17 Mar 2023 13:36:14	U	SMPL
CCV88 - 1	CCV88	23160	5.0565	-	7470A	PPB	17 Mar 2023 13:38:30	U	SMPL
CCB88 - 1	CCB88	46	-0.0054	-	7470A	PPB	17 Mar 2023 13:40:45	U	SMPL
O1907-06 - 1	33961	85	0.0032	-	7470A	PPB	17 Mar 2023 13:43:01	U	SMPL
O1907-07 - 1	33962	108	0.0082	-	7470A	PPB	17 Mar 2023 13:45:17	U	SMPL
O1908-02 - 1	HILLBOR-DRUMS-0314	110	0.0086	-	7470A	PPB	17 Mar 2023 13:47:33	U	SMPL
O1921-01 - 1	COMP-1	107	0.0080	-	7470A	PPB	17 Mar 2023 13:49:49	U	SMPL
O1921-02 - 1	RT3249	88	0.0038	-	7470A	PPB	17 Mar 2023 13:52:05	U	SMPL
O1921-02DUP - 1	RT3249DUP	117	0.0102	-	7470A	PPB	17 Mar 2023 13:54:22	U	SMPL
O1921-02MS - 1	RT3249MS	14998	3.2690	-	7470A	PPB	17 Mar 2023 13:56:39	U	SMPL
O1921-02MSD - 1	RT3249MSD	15572	3.3948	-	7470A	PPB	17 Mar 2023 13:58:56	U	SMPL
CCV89 - 1	CCV89	23990	5.2383	-	7470A	PPB	17 Mar 2023 14:01:13	U	SMPL
CCB89 - 1	CCB89	43	-0.0060	-	7470A	PPB	17 Mar 2023 14:03:29	U	SMPL
PB151425TB - 1	PB151425TB	104	0.0073	-	7470A	PPB	17 Mar 2023 14:05:46	U	SMPL
O1921-02 - 1	SD-5	11	-0.0130	-	7470A	PPB	17 Mar 2023 14:08:02	U	SMPL
O1921-02 - 1	SD-5	21418	4.6750	-	7470A	PPB	17 Mar 2023 14:10:19	U	SMPL
CCV90 - 1	CCV90	22767	4.9704	-	7470A	PPB	17 Mar 2023 14:12:35	U	SMPL
CCB90 - 1	CCB90	69	-0.0003	-	7470A	PPB	17 Mar 2023 14:14:51	U	SMPL