



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124546

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/17/2023	13:25
2	CAL2	1	Water	NA	NA	20.7	7.00	03/17/2023	13:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/17/2023	13:27
4	ICV	1	Water	NA	NA	20.8	6.99	03/17/2023	13:30
5	CCV1	1	Water	NA	NA	20.8	2.01	03/17/2023	13:33
6	O1975-01	1	Solid	20.02	20	22.5	6.38	03/17/2023	13:35
7	O1975-04	1	Solid	20.03	20	22.7	6.82	03/17/2023	13:39
8	O1975-04DUP	1	Solid	20.04	20	22.8	6.83	03/17/2023	13:40
9	CCV2	1	Water	NA	NA	20.7	12.03	03/17/2023	13:42

WORKLIST(Hardcopy Internal Chain)

VB 124546

WorkList Name : ph o1975

WorkList ID : 168247

Department : Wet-Chemistry

Date : 03-17-2023 13:06:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/16/2023	Solid	O1975-01	pH	Cool 4 deg C	H2MA01	J11	CHAMBER-B-72-WEST	03/16/2023	9045D
03/16/2023	Solid	O1975-04	pH	Cool 4 deg C	H2MA01	J11	CHAMBER-C-72-WEST	03/16/2023	9045D

Date/Time 03-17-23 13:20

Raw Sample Received by: 101000

Raw Sample Relinquished by: JTCM

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03-17-23 15:00

JTCM

JTCM