



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124551

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/17/2023	09:40
2	O1946-02		1	14	100.00	>212.0	>212.0	03/17/2023	10:10
3	O1946-03		1	12	100.00	>212.0	>212.0	03/17/2023	10:40
4	O1946-03DUP		1	13	100.00	>212.0	>212.0	03/17/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124551

WorkList Name : fp-3-17

WorkList ID : 168222

Department : Wet-Chemistry

Date : 03-17-2023 08:17:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/23/2023	Water	O1946-02	Flash Point	Cool 4 deg C	PSEG03	J11	NWB-1895	03/16/2023	1010B
03/23/2023	Water	O1946-03	Flash Point	Cool 4 deg C	PSEG03	J11	NWB-1896	03/16/2023	1010B

Date/Time 03/17/2023 09:30
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: PR (CWG)

Date/Time 03/17/2023 11:45
 Raw Sample Received by: PR (CWG)
 Raw Sample Relinquished by: RM CWG