



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124554

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/17/2023	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	03/17/2023	16:01
3	CAL3	1	Water	NA	NA	20.7	10.03	03/17/2023	16:05
4	ICV	1	Water	NA	NA	20.7	6.99	03/17/2023	16:10
5	CCV1	1	Water	NA	NA	20.7	2.01	03/17/2023	16:15
6	O1998-02	1	Solid	20.02	20	22.3	8.76	03/17/2023	16:20
7	O1998-04	1	Solid	20.03	20	22.1	8.72	03/17/2023	16:22
8	O1998-06	1	Solid	20.04	20	22.2	10.08	03/17/2023	16:25
9	O1998-08	1	Solid	20.01	20	22.3	8.85	03/17/2023	16:30
10	O2001-02	1	Solid	20.03	20	23.9	6.99	03/17/2023	16:35
11	O2001-05	1	Solid	20.04	20	23.4	6.88	03/17/2023	16:40
12	O2001-08	1	Solid	20.02	20	23.2	6.89	03/17/2023	16:45
13	O2002-02	1	Solid	20.04	20	23.2	7.58	03/17/2023	16:50
14	O2002-02DUP	1	Solid	20.02	20	23.3	7.59	03/17/2023	16:51
15	CCV2	1	Water	NA	NA	20.7	12.03	03/17/2023	17:00

WORKLIST(Hardcopy Internal Chain)

124554

WorkList Name : CORROSIONITY O200

WorkList ID : 168249

Department : Wet-Chemistry

Date : 03-17-2023 15:32:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/17/2023	Solid	O1998-02	Corrosivity	Cool 4 deg C	TULL02	J11	BRAOD-CHANNEL-SUBSTATION-TP1	03/17/2023	9045D
03/17/2023	Solid	O1998-04	Corrosivity	Cool 4 deg C	TULL02	J11	BRAOD-CHANNEL-SUBSTATION-TP2	03/17/2023	9045D
03/17/2023	Solid	O1998-06	Corrosivity	Cool 4 deg C	TULL02	J11	10TH-ST-SUBSTATION-TP1	03/17/2023	9045D
03/17/2023	Solid	O1998-08	Corrosivity	Cool 4 deg C	TULL02	J11	10TH-ST-SUBSTATION-TP2	03/17/2023	9045D
03/17/2023	Solid	O2001-02	Corrosivity	Cool 4 deg C	PSEG03	J11	IRV-1	03/17/2023	9045D
03/17/2023	Solid	O2001-05	Corrosivity	Cool 4 deg C	PSEG03	J11	IRV-2	03/17/2023	9045D
03/17/2023	Solid	O2001-08	Corrosivity	Cool 4 deg C	PSEG03	J11	IRV-3	03/17/2023	9045D
03/16/2023	Solid	O2002-02	Corrosivity	Cool 4 deg C	PSEG03	J11	031023-3	03/16/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03.17.23

15:45

10/00/1

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03.17.23

14:35

2.6.50m