



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 3/20/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:00

In Date: 03/18/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:00

Out Date: 03/19/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB124560

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1977-01	EYG87	1	1.14	8.58	9.72	7.95	79.4	
O1977-02	EYG88	2	1.15	8.78	9.93	8.39	82.5	
O1977-03	EYG89	3	1.15	8.82	9.97	8.6	84.5	
O1977-04	EYG90	4	1.13	8.70	9.83	8.66	86.6	
O1977-05	EYG91	5	1.16	8.78	9.94	9.00	89.3	
O1977-06	EYG92	6	1.19	8.69	9.88	8.18	80.4	
O1977-07	EYG93	7	1.14	8.82	9.96	8.81	87.0	
O1977-08	EYG94	8	1.13	8.60	9.73	8.02	80.1	
O1977-09	EYG95	9	1.1	8.64	9.74	8.19	82.1	
O1977-10	EYG96	10	1.14	8.83	9.97	8.22	80.2	
O1977-11	EYG97	11	1.12	8.46	9.58	8.12	82.7	
O1977-12	EYG99	12	1.19	8.59	9.78	8.15	81.0	
O1977-13	EYGA0	13	1.15	8.80	9.95	7.92	76.9	
O1977-14	EYGA1	14	1.1	8.75	9.85	7.79	76.5	
O1977-15	EYGA2	15	1.17	8.51	9.68	8.69	88.4	
O1977-16	EYGA2MS	16	1.17	8.51	9.68	8.69	88.4	
O1977-17	EYGA2MSD	17	1.17	8.51	9.68	8.69	88.4	
O1977-18	EYGA3	18	1.15	8.60	9.75	8.23	82.3	
O1977-19	EYGA4	19	1.16	8.81	9.97	8.07	78.4	
O1977-20	EYGA5	20	1.13	8.53	9.66	7.71	77.1	
O1977-21	EYGA6	21	1.13	8.44	9.57	7.65	77.3	
O1977-22	EYGA7	22	1.16	8.81	9.97	8.77	86.4	
O1977-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$