



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 3/20/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:35

In Date: 03/18/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 03/19/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB124562

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1985-01	EYFX5	1	1.13	8.75	9.88	8.43	83.4	
O1985-02	EYFX5MS	2	1.13	8.75	9.88	8.43	83.4	
O1985-03	EYFX5MSD	3	1.13	8.75	9.88	8.43	83.4	
O1985-04	EYFX6	4	1.1	8.58	9.68	8.21	82.9	
O1985-05	EYFX7	5	1.17	8.81	9.98	8.37	81.7	
O1985-06	EYFX8	6	1.16	8.48	9.64	8.12	82.1	
O1985-07	EYFX9	7	1.15	8.80	9.95	8.34	81.7	
O1985-08	EYFY0	8	1.14	8.68	9.82	8.57	85.6	
O1985-09	EYFY1	9	1.15	8.82	9.97	9.26	92.0	
O1985-10	EYFY2	10	1.16	8.81	9.97	8.73	85.9	
O1985-11	EYFY3	11	1.13	8.47	9.6	8.13	82.6	
O1985-12	EYFY4	12	1.18	8.52	9.7	8.21	82.5	
O1985-13	EYFY5	13	1.18	8.63	9.81	8.54	85.3	
O1985-14	EYFY6	14	1.17	8.77	9.94	8.74	86.3	
O1985-15	EYFY7	15	1.17	8.49	9.66	8.58	87.3	
O1985-16	EYFY8	16	1.18	8.45	9.63	8.63	88.2	
O1985-17	EYFY9	17	1.18	8.72	9.9	8.00	78.2	
O1985-18	EYFZ0	18	1.18	8.57	9.75	8.5	85.4	
O1985-19	EYFZ1	19	1.16	8.81	9.97	9.17	90.9	
O1985-20	EYFZ2	20	1.17	8.81	9.98	8.8	86.6	
O1985-21	EYFZ3	21	1.15	8.40	9.55	8.36	85.8	
O1985-22	EYFZ4	22	1.14	8.84	9.98	8.4	82.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$