


LB124576 Instrument ID: CV1

7470A

μ Abs.:

42261



A= 0.0000e+000

B= 2.3868e-004

C= 9.0970e-002

Rho= 0.9958476

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.094	0.094	11	0.000	11	0			
0.2	0.200	0.163	-0.037	301	0.0 %	301				
2.5	2.500	2.962	0.462	12029	0.0 %	12029				
5.0	5.000	4.352	-0.648	17854	0.0 %	17854				
7.5	7.500	7.451	-0.049	30836	0.0 %	30836				
10.0	10.000	10.178	0.178	42261	0.0 %	42261				

4/20/23

LB124576 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Mar 2023 12:30:51					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7470A	PPB	20 Mar 2023 13:44:29	S	Std
0.2 - 1	5.2	301	-	0.2000	7470A	PPB	20 Mar 2023 13:46:46	S	Std
2.5 - 1	52.5	12029	-	2.5000	7470A	PPB	20 Mar 2023 13:49:04	S	Std
5.0 - 1	55.0	17854	-	5.0000	7470A	PPB	20 Mar 2023 13:51:21	S	Std
7.5 - 1	57.5	30836	-	7.5000	7470A	PPB	20 Mar 2023 13:53:38	S	Std
10.0 - 1	510	42261	-	10.0000	7470A	PPB	20 Mar 2023 14:30:18	S	Std
ICV99 - 1	ICV99	15964	3.9013	-	7470A	PPB	20 Mar 2023 14:42:00	U	SMPL
ICB99 - 1	ICB99	-27	0.0845	-	7470A	PPB	20 Mar 2023 14:44:15	U	SMPL
CCV001 - 1	CCV001	19933	4.8487	-	7470A	PPB	20 Mar 2023 14:49:37	U	SMPL
CCB001 - 1	CCB001	-103	0.0664	-	7470A	PPB	20 Mar 2023 14:51:52	U	SMPL
CRA - 1	CRA	590	0.2318	-	7470A	PPB	20 Mar 2023 14:54:09	U	SMPL
HighStd - 1	HighStd	41723	10.0496	-	7470A	PPB	20 Mar 2023 14:56:24	U	SMPL
ChkStd - 1	ChkStd	28103	6.7987	-	7470A	PPB	20 Mar 2023 14:58:39	U	SMPL
PB151500BL - 1	PBW	-170	0.0504	-	7470A	PPB	20 Mar 2023 15:00:58	U	SMPL
PB151500BS - 1	LCSW	17522	4.2732	-	7470A	PPB	20 Mar 2023 15:03:15	U	SMPL
O1928-02 - 1	031423-1	-65	0.0755	-	7470A	PPB	20 Mar 2023 15:05:32	U	SMPL
O1932-04 - 1	CE-86B-031423	45	0.1017	-	7470A	PPB	20 Mar 2023 15:07:49	U	SMPL
O1932-08 - 1	CE-81-031423	39	0.1003	-	7470A	PPB	20 Mar 2023 15:10:06	U	SMPL
O1932-12 - 1	CE-43-031423	113	0.1179	-	7470A	PPB	20 Mar 2023 15:12:23	U	SMPL
O1932-16 - 1	JPP-6A-031423	71	0.1079	-	7470A	PPB	20 Mar 2023 15:14:39	U	SMPL
CCV198 - 1	CCV198	19120	4.6546	-	7470A	PPB	20 Mar 2023 15:16:54	U	SMPL
CCB198 - 1	CCB198	-74	0.0733	-	7470A	PPB	20 Mar 2023 15:19:10	U	SMPL
O1932-20 - 1	JPP-6C-031423	16	0.0948	-	7470A	PPB	20 Mar 2023 15:21:27	U	SMPL
O1939-01 - 1	D2945	-14	0.0876	-	7470A	PPB	20 Mar 2023 15:23:43	U	SMPL
O1939-02 - 1	D2946	106	0.1163	-	7470A	PPB	20 Mar 2023 15:25:58	U	SMPL
O1939-03 - 1	D2943	39	0.1003	-	7470A	PPB	20 Mar 2023 15:28:14	U	SMPL
O1942-01 - 1	350	33	0.0988	-	7470A	PPB	20 Mar 2023 15:30:30	U	SMPL
O1954-02 - 1	SV-TP1-0316	40	0.1005	-	7470A	PPB	20 Mar 2023 15:32:47	U	SMPL
O1954-04 - 1	SV-TP2-0316	28	0.0977	-	7470A	PPB	20 Mar 2023 15:35:04	U	SMPL
O1954-06 - 1	SV-TP3-0316	45	0.1017	-	7470A	PPB	20 Mar 2023 15:37:21	U	SMPL
O1954-08 - 1	SV-TP4-0316	44	0.1015	-	7470A	PPB	20 Mar 2023 15:39:38	U	SMPL
O1954-10 - 1	SV-TP5-0316	50	0.1029	-	7470A	PPB	20 Mar 2023 15:41:54	U	SMPL
CCV199 - 1	CCV199	20371	4.9532	-	7470A	PPB	20 Mar 2023 15:44:09	U	SMPL
CCB199 - 1	CCB199	-93	0.0688	-	7470A	PPB	20 Mar 2023 15:46:25	U	SMPL
O1954-10DUP - 1	SV-TP5-0316DUP	29	0.0979	-	7470A	PPB	20 Mar 2023 15:48:42	U	SMPL
O1954-10MS - 1	SV-TP5-0316MS	14018	3.4368	-	7470A	PPB	20 Mar 2023 15:50:58	U	SMPL
O1954-10MSD - 1	SV-TP5-0316MSD	14153	3.4691	-	7470A	PPB	20 Mar 2023 15:53:14	U	SMPL
PB151501BL - 1	PBW	-45	0.0802	-	7470A	PPB	20 Mar 2023 15:55:31	U	SMPL
PB151501BS - 1	LCSW	16732	4.0846	-	7470A	PPB	20 Mar 2023 15:57:48	U	SMPL
O1967-01DUP - 1	3X1DUP	619	0.2387	-	7470A	PPB	20 Mar 2023 16:02:21	U	SMPL
O1967-01MS - 1	3X1MS	14235	3.4886	-	7470A	PPB	20 Mar 2023 16:04:40	U	SMPL
O1967-01MSD - 1	3X1MSD	14402	3.5285	-	7470A	PPB	20 Mar 2023 16:06:57	U	SMPL
O1967-02 - 1	3X2	706	0.2595	-	7470A	PPB	20 Mar 2023 16:09:13	U	SMPL
CCV004 - 1	CCV004	19454	4.7343	-	7470A	PPB	20 Mar 2023 16:09:13	U	SMPL
CCB004 - 1	CCB004	-84	0.0709	-	7470A	PPB	20 Mar 2023 16:11:30	U	SMPL
O1967-03 - 1	3X3	1326	0.4075	-	7470A	PPB	20 Mar 2023 16:13:46	U	SMPL
O1967-04 - 1	3X4	794	0.2805	-	7470A	PPB	20 Mar 2023 16:16:03	U	SMPL
O1967-05 - 1	3X5	-28	0.0843	-	7470A	PPB	20 Mar 2023 16:18:19	U	SMPL
O1967-06 - 1	3X6	-12	0.0881	-	7470A	PPB	20 Mar 2023 16:20:35	U	SMPL
O1967-07 - 1	3X7	181	0.1342	-	7470A	PPB	20 Mar 2023 16:22:51	U	SMPL
O1967-08 - 1	3X8	80	0.1101	-	7470A	PPB	20 Mar 2023 16:25:07	U	SMPL
O1967-09 - 1	3X9	115	0.1184	-	7470A	PPB	20 Mar 2023 16:27:24	U	SMPL
O1967-10 - 1	3X10	0	0.0910	-	7470A	PPB	20 Mar 2023 16:29:40	U	SMPL
B151506BL - 1	PBW	45	0.1017	-	7470A	PPB	20 Mar 2023 16:31:58	U	SMPL
B151506BS - 1	LCSW	19494	4.7439	-	7470A	PPB	20 Mar 2023 16:34:15	U	SMPL
CV005 - 1	CCV005	21137	5.1360	-	7470A	PPB	20 Mar 2023 16:36:31	U	SMPL
CB005 - 1	CCB005	-88	0.0700	-	7470A	PPB	20 Mar 2023 16:40:17	U	SMPL
O1939-06 - 1	D2958	51	0.1031	-	7470A	PPB	20 Mar 2023 16:42:32	U	SMPL
O1942-02 - 1	349	40	0.1005	-	7470A	PPB	20 Mar 2023 16:44:49	U	SMPL
O1942-02DUP - 1	349DUP	40	0.1005	-	7470A	PPB	20 Mar 2023 16:47:05	U	SMPL
O1942-02MS - 1	349MS	14733	3.6075	-	7470A	PPB	20 Mar 2023 16:49:21	U	SMPL
O1942-02MSD - 1	349MSD	13746	3.3719	-	7470A	PPB	20 Mar 2023 16:51:37	U	SMPL
O1954-10LX5 - 1		0	0.0910	-	7470A	PPB	20 Mar 2023 16:53:53	U	SMPL
O1967-01LX5 - 1		35	0.0993	-	7470A	PPB	20 Mar 2023 16:56:09	U	SMPL
O1942-02LX5 - 1		22	0.0962	-	7470A	PPB	20 Mar 2023 16:58:26	U	SMPL
CV006 - 1	CCV006	21187	5.1480	-	7470A	PPB	20 Mar 2023 17:00:43	U	SMPL
CB006 - 1	CCB006	-86	0.0704	-	7470A	PPB	20 Mar 2023 17:07:33	U	SMPL
B151467TB - 1		-3	0.0903	-	7470A	PPB	20 Mar 2023 17:09:49	U	SMPL
O1967-01 - 1	3X1	988	0.3268	-	7470A	PPB	20 Mar 2023 17:18:56	U	SMPL
B151465TB - 1		-143	0.0568	-	7470A	PPB	20 Mar 2023 17:21:12	U	SMPL
B151497TB - 1		-10	0.0886	-	7470A	PPB	20 Mar 2023 17:23:28	U	SMPL
CV007 - 1	CCV007	19084	4.6460	-	7470A	PPB	20 Mar 2023 17:25:46	U	SMPL
CB007 - 1	CCB007	-106	0.0657	-	7470A	PPB	20 Mar 2023 17:31:32	U	SMPL
					7470A	PPB	20 Mar 2023 17:33:48	U	SMPL