



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124593

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/20/2023	17:25
2	CAL2	1	Water	NA	NA	20.7	6.99	03/20/2023	17:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/20/2023	17:28
4	ICV	1	Water	NA	NA	20.8	6.98	03/20/2023	17:30
5	CCV1	1	Water	NA	NA	20.7	2.01	03/20/2023	17:33
6	O1981-01	1	Water	NA	NA	20.3	4.88	03/20/2023	17:35
7	O2013-08	1	Water	NA	NA	20.7	6.68	03/20/2023	17:39
8	O2020-41	1	Water	NA	NA	20.6	12.12	03/20/2023	17:45
9	O2020-41DUP	1	Water	NA	NA	20.7	12.13	03/20/2023	17:46
10	CCV2	1	Water	NA	NA	20.7	12.03	03/20/2023	17:49

WORKLIST(Hardcopy Internal Chain)

VB 124593

WorkList Name : PH I2013 W

WorkList ID : 168322

Department : Wet-Chemistry

Date : 03-20-2023 15:57:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/28/2023	Water	O1981-01	pH	Cool 4 deg C	STER06	M23	POOL-WATER	02/28/2023	9040C
03/20/2023	Water	O2013-08	pH	Cool 4 deg C	PSEG03	J11	215-217	03/20/2023	9040C
03/20/2023	Water	O2020-41	pH	Cool 4 deg C	PSEG03	K11	FRAC-TANK-A821	03/20/2023	9040C

03.20.23

17:20

Date/Time
Raw Sample Received by: JO (COC)
Raw Sample Relinquished by: OP SM

Date/Time
Raw Sample Received by: OP SM
Raw Sample Relinquished by: JO (COC)