


LB124597 INSTRUMENT? CV,

7470A

Linear

 μ Abs.:

39077



A= 0.0000e+000

B= 2.5592e-004

C= -6.6483e-002

Rho= 0.9997940

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/D
0.0	0.000	-0.034	-0.034	127	0.000	127	0				1
0.2	0.200	0.144	-0.056	824	0.0 %	824					-28
2.5	2.500	2.529	0.029	10143	0.0 %	10143					12
5.0	5.000	5.153	0.153	20396	0.0 %	20396					3
7.5	7.500	7.473	-0.027	29461	0.0 %	29461					-4
10.0	10.000	9.934	-0.066	39077	0.0 %	39077					1

LB124597, INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Mar 2023 12:55:49

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	500	127	-	0.0000	7470A	PPB	21 Mar 2023 12:56:11	S	Std
0.2 - 1	502	824	-	0.2000	7470A	PPB	21 Mar 2023 12:58:29	S	Std
2.5 - 1	525	10143	-	2.5000	7470A	PPB	21 Mar 2023 13:00:46	S	Std
5.0 - 1	550	20396	-	5.0000	7470A	PPB	21 Mar 2023 13:03:03	S	Std
7.5 - 1	575	29461	-	7.5000	7470A	PPB	21 Mar 2023 13:05:20	S	Std
10.0 - 1	5100	39077	-	10.0000	7470A	PPB	21 Mar 2023 13:07:37	S	Std
ICV06 - 1	ICV06	17413	4.3898	-	7470A	PPB	21 Mar 2023 13:14:30	U	SMPL
ICB06 - 1	ICB06	18	-0.0619	-	7470A	PPB	21 Mar 2023 13:16:45	U	SMPL
CCV18 - 1	CCV18	19047	4.8080	-	7470A	PPB	21 Mar 2023 13:19:01	U	SMPL
CCB18 - 1	CCB18	13	-0.0632	-	7470A	PPB	21 Mar 2023 13:21:16	U	SMPL
CRA - 1	CRA	1052	0.2027	-	7470A	PPB	21 Mar 2023 13:23:32	U	SMPL
HighStd - 1	HighStd	37866	9.6241	-	7470A	PPB	21 Mar 2023 13:25:48	U	SMPL
ChkStd - 1	ChkStd	25491	6.4571	-	7470A	PPB	21 Mar 2023 13:28:03	U	SMPL
PB151569BL - 1	PBW	-75	-0.0857	-	7470A	PPB	21 Mar 2023 13:30:21	U	SMPL
PB151569BS - 1	LCSW	15868	3.9944	-	7470A	PPB	21 Mar 2023 13:32:38	U	SMPL
O1963-02 - 1	NYE-SOIL-0316	7	-0.0647	-	7470A	PPB	21 Mar 2023 13:34:55	U	SMPL
O1964-02 - 1	UNION-2	49	-0.0539	-	7470A	PPB	21 Mar 2023 13:37:12	U	SMPL
O1964-05 - 1	MAPLE-4	73	-0.0478	-	7470A	PPB	21 Mar 2023 13:39:30	U	SMPL
O1964-08 - 1	MAPLE-5	66	-0.0496	-	7470A	PPB	21 Mar 2023 13:41:47	U	SMPL
O1964-11 - 1	MAPLE-6	97	-0.0417	-	7470A	PPB	21 Mar 2023 13:44:02	U	SMPL
CCV19 - 1	CCV19	17972	4.5329	-	7470A	PPB	21 Mar 2023 13:46:17	U	SMPL
CCB19 - 1	CCB19	24	-0.0603	-	7470A	PPB	21 Mar 2023 13:48:33	U	SMPL
O1968-01 - 1	149TH-ST-DRUM-13-POLY	29	-0.0591	-	7470A	PPB	21 Mar 2023 13:50:49	U	SMPL
O1968-02 - 1	149TH-ST-DRUM-14-PPE	61	-0.0509	-	7470A	PPB	21 Mar 2023 13:53:04	U	SMPL
O1974-02 - 1	SHELL-ROAD-1	80	-0.0480	-	7470A	PPB	21 Mar 2023 13:55:20	U	SMPL
O1974-05 - 1	SHELL-ROAD-2	68	-0.0491	-	7470A	PPB	21 Mar 2023 13:57:36	U	SMPL
O1975-02 - 1	CHAMBER-B-72-WEST	73	-0.0478	-	7470A	PPB	21 Mar 2023 13:59:52	U	SMPL
O1975-05 - 1	CHAMBER-C-72-WEST	110	-0.0383	-	7470A	PPB	21 Mar 2023 14:02:08	U	SMPL
O1998-02 - 1	BROAD-CHANNEL-SUBSTATION-TP1	69	-0.0488	-	7470A	PPB	21 Mar 2023 14:04:25	U	SMPL
O1998-04 - 1	BROAD-CHANNEL-SUBSTATION-TP2	75	-0.0473	-	7470A	PPB	21 Mar 2023 14:06:42	U	SMPL
O1998-06 - 1	10TH-ST-SUBSTATION-TP1	66	-0.0496	-	7470A	PPB	21 Mar 2023 14:09:00	U	SMPL
O1998-08 - 1	10TH-ST-SUBSTATION-TP2	125	-0.0345	-	7470A	PPB	21 Mar 2023 14:11:15	U	SMPL
CCV20 - 1	CCV20	18462	4.6583	-	7470A	PPB	21 Mar 2023 14:13:31	U	SMPL
CCB20 - 1	CCB20	-19	-0.0713	-	7470A	PPB	21 Mar 2023 14:15:47	U	SMPL
O2001-02 - 1	IRV-1	62	-0.0506	-	7470A	PPB	21 Mar 2023 14:18:04	U	SMPL
O2001-05 - 1	IRV-2	51	-0.0534	-	7470A	PPB	21 Mar 2023 14:20:19	U	SMPL
O2001-08 - 1	IRV-3	34	-0.0578	-	7470A	PPB	21 Mar 2023 14:22:35	U	SMPL
O2001-08DUP - 1	IRV-3DUP	-20	-0.0716	-	7470A	PPB	21 Mar 2023 14:24:51	U	SMPL
O2001-08MS - 1	IRV-3MS	15812	3.9801	-	7470A	PPB	21 Mar 2023 14:27:07	U	SMPL
O2001-08MSD - 1	IRV-3MSD	16117	4.0581	-	7470A	PPB	21 Mar 2023 14:29:24	U	SMPL
PB151584BL - 1	PBW	-4	-0.0675	-	7470A	PPB	21 Mar 2023 14:31:41	U	SMPL
O2002-02 - 1	031023-3	26	-0.0598	-	7470A	PPB	21 Mar 2023 14:36:15	U	SMPL
O2016-02 - 1	032023-1	119	-0.0360	-	7470A	PPB	21 Mar 2023 14:38:32	U	SMPL
CCV21 - 1	CCV21	19670	4.9674	-	7470A	PPB	21 Mar 2023 14:40:47	U	SMPL
CCB21 - 1	CCB21	-1	-0.0667	-	7470A	PPB	21 Mar 2023 14:43:03	U	SMPL
O2018-02 - 1	DAHMER-ROCKY-SOIL	99	-0.0411	-	7470A	PPB	21 Mar 2023 14:45:21	U	SMPL
O2018-02DUP - 1	DAHMER-ROCKY-SOILDUP	91	-0.0432	-	7470A	PPB	21 Mar 2023 14:47:36	U	SMPL
O2018-02MS - 1	DAHMER-ROCKY-SOILMS	12587	3.1647	-	7470A	PPB	21 Mar 2023 14:49:52	U	SMPL
O2018-02MSD - 1	DAHMER-ROCKY-SOILMSD	11607	2.9039	-	7470A	PPB	21 Mar 2023 14:52:08	U	SMPL
O2001-08LX5 - 1		-58	-0.0813	-	7470A	PPB	21 Mar 2023 14:54:24	U	SMPL
O2001-08A - 1		16305	4.1062	-	7470A	PPB	21 Mar 2023 14:56:41	U	SMPL
O2018-02LX5 - 1		-56	-0.0808	-	7470A	PPB	21 Mar 2023 14:58:57	U	SMPL
O2018-02A - 1		16327	4.1119	-	7470A	PPB	21 Mar 2023 15:01:15	U	SMPL
PB151496TB - 1		-25	-0.0729	-	7470A	PPB	21 Mar 2023 15:03:32	U	SMPL
PB151560TB - 1		86	-0.0445	-	7470A	PPB	21 Mar 2023 15:05:49	U	SMPL
CCV22 - 1	CCV22	20181	5.0982	-	7470A	PPB	21 Mar 2023 15:08:04	U	SMPL
CCB22 - 1	CCB22	-23	-0.0724	-	7470A	PPB	21 Mar 2023 15:10:20	U	SMPL
PB151584BS - 1		16106	4.0553	-	7470A	PPB	21 Mar 2023 15:12:37	U	SMPL
CCV23 - 1	CCV23	18802	4.7453	-	7470A	PPB	21 Mar 2023 15:14:52	U	SMPL
CCB23 - 1	CCB23	-39	-0.0765	-	7470A	PPB	21 Mar 2023 15:17:08	U	SMPL

3/21/23