



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124600

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/21/2023	08:50
2	O2013-01		1	14	100.00	>212.0	>212.0	03/21/2023	09:20
3	O2013-01DUP		1	14	100.00	>212.0	>212.0	03/21/2023	09:50
4	O2013-02		1	12	100.00	>212.0	>212.0	03/21/2023	10:20
5	O2013-03		1	13	100.00	>212.0	>212.0	03/21/2023	10:50
6	O2013-04		1	14	100.00	>212.0	>212.0	03/21/2023	11:20
7	O2013-08		1	14	100.00	>212.0	>212.0	03/21/2023	11:50
8	O2014-01		1	13	100.00	>212.0	>212.0	03/21/2023	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 124600

WorkList Name : fp-3-20

WorkList ID : 168295

Department : Wet-Chemistry

Date : 03-20-2023 08:30:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/27/2023	Water	O2013-01	Flash Point	Cool 4 deg C	PSEG03	J11	600	03/20/2023	1010B
03/27/2023	Water	O2013-02	Flash Point	Cool 4 deg C	PSEG03	J11	31993	03/20/2023	1010B
03/27/2023	Water	O2013-03	Flash Point	Cool 4 deg C	PSEG03	J11	218	03/20/2023	1010B
03/27/2023	Water	O2013-04	Flash Point	Cool 4 deg C	PSEG03	J11	219	03/20/2023	1010B
03/27/2023	Water	O2013-08	Flash Point	Cool 4 deg C	PSEG03	J11	215-217	03/20/2023	1010B
03/27/2023	Water	O2014-01	Flash Point	Cool 4 deg C	PSEG03	J11	1711	03/20/2023	1010B

Date/Time 03/21/2023 08:40
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JN (RM)

Date/Time 03/21/2023 13:00
 Raw Sample Received by: JN (RM)
 Raw Sample Relinquished by: RM CWC