



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124606

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	03/21/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	03/21/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/21/2023	16:30
4	ICV	1	Water	NA	NA	20.8	6.99	03/21/2023	16:35
5	CCV1	1	Water	NA	NA	20.6	2.01	03/21/2023	16:39
6	O2022-01	1	Solid	20.02	20	22.3	9.83	03/21/2023	16:45
7	O2022-01DUP	1	Solid	20.03	20	22.4	9.84	03/21/2023	16:46
8	CCV2	1	Water	NA	NA	20.7	12.02	03/21/2023	16:50

WORKLIST(Hardcopy Internal Chain)

VB124606

WorkList Name : ph o2022

WorkList ID : 168368

Department : Wet-Chemistry

Date : 03-21-2023 16:15:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2023	Solid	O2022-01	pH	A	Cool 4 deg C	PSEG05 L11	AU-05-32123	03/21/2023	9045D

Date/Time 03/21/23 16:20
 Raw Sample Received by: 70 (WC1)
 Raw Sample Relinquished by: JT (9m)

Date/Time 03/21/23 18:00
 Raw Sample Received by: JT (9m)
 Raw Sample Relinquished by: 70 (WC1)