



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124607

Slope : 98.6

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	03/21/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	03/21/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/21/2023	16:30
4	ICV	1	Water	NA	NA	20.8	6.99	03/21/2023	16:35
5	CCV1	1	Water	NA	NA	20.7	2.02	03/21/2023	17:00
6	O2027-01	1	Water	NA	NA	20.7	7.14	03/21/2023	17:05
7	O2027-01DUP	1	Water	NA	NA	20.8	7.15	03/21/2023	17:06
8	O2027-03	1	Water	NA	NA	20.1	7.39	03/21/2023	17:10
9	O2027-05	1	Water	NA	NA	20.6	5.75	03/21/2023	17:15
10	CCV2	1	Water	NA	NA	20.7	12.03	03/21/2023	17:20

2/3/24607

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2027 w

WorkList ID : 168369

Department : Wet-Chemistry

Date : 03-21-2023 16:16:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2023	Water	O2027-01	pH	Cool 4 deg C	PSEG03	L11	COMP-1	03/21/2023	9040C
03/21/2023	Water	O2027-03	pH	Cool 4 deg C	PSEG03	L11	COMP-2	03/21/2023	9040C
03/21/2023	Water	O2027-05	pH	Cool 4 deg C	PSEG03	L11	COMP-3	03/21/2023	9040C

Date/Time 03/21/23 16:25
Raw Sample Received by: 70 (WC)
Raw Sample Relinquished by: JTC (Sm)

Date/Time 03/21/23 18:00
Raw Sample Received by: JTC (Sm)
Raw Sample Relinquished by: 70 (WC)