


LB124611 INSTRUMENT ; CV1

7471B

Linear

μ Abs. :

50883



A= 0.0000e+000

B= 1.9426e-004

C= 4.0057e-003

Rho= 0.9996224

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	48	0.000	48				
0.2	0.200	0.228	0.028	1151	0.0 %	1151				
2.5	2.500	2.422	-0.078	12445	0.0 %	12445				
5.0	5.000	4.950	-0.050	25458	0.0 %	25458				
7.5	7.500	7.699	0.199	39612	0.0 %	39612				
10.0	10.000	9.889	-0.111	50883	0.0 %	50883				

1.0
14
-3
-1
0
-1

LB124611 INSTRUMENT ID: CV1

Method: 7471B	Operator: Admin	Date of Analysis: 21 Mar 2023 18:04:30							
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	48	-	0.000	7471B	PPB	21 Mar 2023 18:04:46	S	Std
0.2 - 1	50.2	1151	-	0.200	7471B	PPB	21 Mar 2023 18:07:03	S	Std
2.5 - 1	52.5	12445	-	2.500	7471B	PPB	21 Mar 2023 18:09:21	S	Std
5.0 - 1	55.0	25458	-	5.000	7471B	PPB	21 Mar 2023 18:11:38	S	Std
7.5 - 1	57.5	39612	-	7.500	7471B	PPB	21 Mar 2023 18:13:55	S	Std
10.0 - 1	510.0	50883	-	10.000	7471B	PPB	21 Mar 2023 18:24:07	S	Std
ICV07 - 1	ICV07	20427	3.9722	-	7471B	PPB	21 Mar 2023 18:32:54	U	SMPL
ICB07 - 1	ICB07	-36	-0.0030	-	7471B	PPB	21 Mar 2023 18:35:09	U	SMPL
CCV24 - 1	CCV24	25251	4.9094	-	7471B	PPB	21 Mar 2023 18:37:25	U	SMPL
CCB24 - 1	CCB24	-48	-0.0053	-	7471B	PPB	21 Mar 2023 18:39:40	U	SMPL
CRA - 1	CRA	1125	0.2226	-	7471B	PPB	21 Mar 2023 18:41:58	U	SMPL
HighStd - 1	HighStd	49516	9.6232	-	7471B	PPB	21 Mar 2023 18:44:13	U	SMPL
ChkStd - 1	ChkStd	33396	6.4916	-	7471B	PPB	21 Mar 2023 18:46:29	U	SMPL
PB151590BL - 1	PB151590BL	-53	-0.0063	-	7471B	PPB	21 Mar 2023 18:48:47	U	SMPL
PB151590BS - 1	PB151590BS	19304	3.7541	-	7471B	PPB	21 Mar 2023 18:51:06	U	SMPL
O2012-01 - 1	B-23-SB01	11956	2.3266	-	7471B	PPB	21 Mar 2023 18:53:22	U	SMPL
O2012-02 - 1	B-23-SB02	1819	0.3574	-	7471B	PPB	21 Mar 2023 18:55:40	U	SMPL
O2016-01 - 1	032023-1	2513	0.4922	-	7471B	PPB	21 Mar 2023 18:57:57	U	SMPL
O2018-01 - 1	DAHMER-ROCKY-SOIL	1413	0.2785	-	7471B	PPB	21 Mar 2023 19:00:15	U	SMPL
O2019-01 - 1	ETGI-352	3612	0.7057	-	7471B	PPB	21 Mar 2023 19:02:30	U	SMPL
CCV25 - 1	CCV25	23872	4.6415	-	7471B	PPB	21 Mar 2023 19:04:45	U	SMPL
CCB25 - 1	CCB25	-46	-0.0049	-	7471B	PPB	21 Mar 2023 19:07:01	U	SMPL
O2020-01 - 1	SP-1	1800	0.3537	-	7471B	PPB	21 Mar 2023 19:09:18	U	SMPL
O2020-11 - 1	SP-2	1626	0.3199	-	7471B	PPB	21 Mar 2023 19:11:33	U	SMPL
O2020-21 - 1	SP-3	355	0.0730	-	7471B	PPB	21 Mar 2023 19:13:49	U	SMPL
O2020-31 - 1	SP-4	569	0.1145	-	7471B	PPB	21 Mar 2023 19:16:05	U	SMPL
O2021-03 - 1	JPP-6E-032023	1754	0.3447	-	7471B	PPB	21 Mar 2023 19:18:21	U	SMPL
O2021-03DUP - 1	JPP-6E-032023DUP	1940	0.3809	-	7471B	PPB	21 Mar 2023 19:20:37	U	SMPL
O2021-03MS - 1	JPP-6E-032023MS	16876	3.2824	-	7471B	PPB	21 Mar 2023 19:22:54	U	SMPL
O2021-03MSD - 1	JPP-6E-032023MSD	16155	3.1423	-	7471B	PPB	21 Mar 2023 19:25:11	U	SMPL
O2021-07 - 1	JPP-6D-032023	18595	3.6163	-	7471B	PPB	21 Mar 2023 19:27:29	U	SMPL
O2021-11 - 1	JPP-46A-032023	31412	6.1062	-	7471B	PPB	21 Mar 2023 19:29:46	U	SMPL
CCV26 - 1	CCV26	25119	4.8837	-	7471B	PPB	21 Mar 2023 19:35:56	U	SMPL
CCB26 - 1	CCB26	-30	-0.0018	-	7471B	PPB	21 Mar 2023 19:38:11	U	SMPL
O2021-15 - 1	JPP-46B-032023	10651	2.0731	-	7471B	PPB	21 Mar 2023 19:40:29	U	SMPL
O2023-01 - 1	TP-11-TEST-PIT-ON-CENTRAL	518	0.1046	-	7471B	PPB	21 Mar 2023 19:42:44	U	SMPL
O2023-03 - 1	TP-12-TEST-PIT-ON-OCEAN	1574	0.3098	-	7471B	PPB	21 Mar 2023 19:45:00	U	SMPL
O2023-05 - 1	TP-13-TEST-PIT-ON-GRAVELLY	3908	0.7632	-	7471B	PPB	21 Mar 2023 19:47:16	U	SMPL
O2021-03LX5 - 1		335	0.0691	-	7471B	PPB	21 Mar 2023 19:49:32	U	SMPL
O2021-03A - 1		19590	3.8096	-	7471B	PPB	21 Mar 2023 19:51:49	U	SMPL
CCV27 - 1	CCV27	23650	4.5983	-	7471B	PPB	21 Mar 2023 19:54:05	U	SMPL
CCB27 - 1	CCB27	-14	0.0013	-	7471B	PPB	21 Mar 2023 19:56:24	U	SMPL
PB151540BL - 1		69	0.0174	-	7471B	PPB	21 Mar 2023 19:58:42	U	SMPL
PB151540BS - 1		18391	3.5767	-	7471B	PPB	21 Mar 2023 20:00:58	U	SMPL
O1924-01 - 1		1429	0.2816	-	7471B	PPB	21 Mar 2023 20:03:14	U	SMPL
O1924-01DUP - 1		1515	0.2983	-	7471B	PPB	21 Mar 2023 20:05:32	U	SMPL
O1924-01LX5 - 1		315	0.0652	-	7471B	PPB	21 Mar 2023 20:07:47	U	SMPL
O1924-01MS - 1		17511	3.4058	-	7471B	PPB	21 Mar 2023 20:10:03	U	SMPL
O1924-01MSD - 1		17432	3.3904	-	7471B	PPB	21 Mar 2023 20:12:19	U	SMPL
O1924-01A - 1		23036	4.4791	-	7471B	PPB	21 Mar 2023 20:14:36	U	SMPL
CCV28 - 1	CCV28	27031	5.2551	-	7471B	PPB	21 Mar 2023 20:19:37	U	SMPL
CCB28 - 1	CCB28	-39	-0.0036	-	7471B	PPB	21 Mar 2023 20:21:53	U	SMPL
O1932-03 - 1		9537	1.8567	-	7471B	PPB	21 Mar 2023 20:24:12	U	SMPL
O1932-03DUP - 1		11659	2.2689	-	7471B	PPB	21 Mar 2023 20:26:29	U	SMPL
O1932-03LX5 - 1		1937	0.3803	-	7471B	PPB	21 Mar 2023 20:32:44	U	SMPL
O1932-03MS - 1		26357	5.1242	-	7471B	PPB	21 Mar 2023 20:35:00	U	SMPL
O1932-03MSD - 1		26175	5.0889	-	7471B	PPB	21 Mar 2023 20:37:15	U	SMPL
O1932-03A - 1		28148	5.4716	-	7471B	PPB	21 Mar 2023 20:49:05	U	SMPL
CCV29 - 1	CCV29	26273	5.1079	-	7471B	PPB	21 Mar 2023 20:51:21	U	SMPL
CCB29 - 1	CCB29	-57	-0.0071	-	7471B	PPB	21 Mar 2023 20:53:38	U	SMPL