



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124619

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/22/2023	12:40
2	CAL2	1	Water	NA	NA	20.7	7.00	03/22/2023	12:41
3	CAL3	1	Water	NA	NA	20.7	10.03	03/22/2023	12:45
4	ICV	1	Water	NA	NA	20.8	7.00	03/22/2023	12:48
5	CCV1	1	Water	NA	NA	20.7	2.02	03/22/2023	12:50
6	O1953-21	1	Water	NA	NA	23.2	8.20	03/22/2023	13:00
7	O2043-01	1	Water	NA	NA	21.7	4.11	03/22/2023	13:05
8	O2043-01DUP	1	Water	NA	NA	21.8	4.12	03/22/2023	13:06
9	CCV2	1	Water	NA	NA	20.7	12.02	03/22/2023	13:15

WORKLIST(Hardcopy Internal Chain)

12/16/23

WorkList Name : PH O2043 W WorkList ID : 168394 Department : Wet-Chemistry Date : 03-22-2023 12:27:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/14/2023	Water	O1953-21	pH	Cool 4 deg C	CHEM02	QA Of	PT-PH-WP	03/14/2023	9040C
03/16/2023	Water	O2043-01	pH	Cool 4 deg C	STER15	L11	POOL-WATER	03/16/2023	9040C

Date/Time 03/22/23 12:27
 Raw Sample Received by: JP/OC
 Raw Sample Relinquished by: SJ (QA office)

Date/Time 03/22/23 14:00
 Raw Sample Received by: SJ (QA office)
 Raw Sample Relinquished by: JP/OC