



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124620

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/22/2023	12:40
2	CAL2	1	Water	NA	NA	20.7	7.00	03/22/2023	12:41
3	CAL3	1	Water	NA	NA	20.7	10.03	03/22/2023	12:45
4	ICV	1	Water	NA	NA	20.8	7.00	03/22/2023	12:48
5	CCV1	1	Water	NA	NA	20.7	2.02	03/22/2023	12:50
6	O1953-20	1	Water	NA	NA	23.2	8.20	03/22/2023	13:00
7	CCV2	1	Water	NA	NA	20.7	12.02	03/22/2023	13:15

WORKLIST(Hardcopy Internal Chain)

12/26/20

WorkList Name : ph o1953 w

WorkList ID : 168395

Department : Wet-Chemistry

Date : 03-22-2023 12:43:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/14/2023	Water	O1953-20	pH	Cool 4 deg C	CHEM02	QA Of	PT-PH-WP	03/14/2023	SM 4500-H B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03/22/23 12:35

78 (QC)

SI (QA-Office)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03/22/23

SI (QA-Office)

78 (QC)