


7470A

Linear

 μ Abs.:

29590



A= 0.0000e+000

B= 3.3843e-004

C= -3.0314e-002

Rho= 0.9999669

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	55	0.000	55	0				5
0.2	0.200	0.209	0.009	707	0.0 %	707					1
2.5	2.500	2.522	0.022	7541	0.0 %	7541					1
5.0	5.000	4.951	-0.049	14720	0.0 %	14720					1
7.5	7.500	7.546	0.046	22386	0.0 %	22386					1
10.0	10.000	9.984	-0.016	29590	0.0 %	29590					0

LB124626 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Mar 2023 16:59:05

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1		55	-	0.0000	7470A	PPB	22 Mar 2023 17:15:30	S	Std
0.2 - 1		707	-	0.2000	7470A	PPB	22 Mar 2023 17:17:48	S	Std
2.5 - 1		7541	-	2.5000	7470A	PPB	22 Mar 2023 17:20:05	S	Std
5.0 - 1		14720	-	5.0000	7470A	PPB	22 Mar 2023 17:22:22	S	Std
7.5 - 1		22386	-	7.5000	7470A	PPB	22 Mar 2023 17:24:39	S	Std
10.0 - 1		29590	-	10.0000	7470A	PPB	22 Mar 2023 17:26:56	S	Std
ICV08 - 1	ICV08	12983	4.3635	-	7470A	PPB	22 Mar 2023 17:33:17	U	SMPL
ICB08 - 1	ICB08	-10	-0.0337	-	7470A	PPB	22 Mar 2023 17:35:32	U	SMPL
CCV30 - 1	CCV30	14665	4.9328	-	7470A	PPB	22 Mar 2023 17:37:48	U	SMPL
CCB30 - 1	CCB30	-17	-0.0361	-	7470A	PPB	22 Mar 2023 17:40:04	U	SMPL
CRA - 1	CRA	645	0.1880	-	7470A	PPB	22 Mar 2023 17:42:20	U	SMPL
HighStd - 1	HighStd	28409	9.5841	-	7470A	PPB	22 Mar 2023 17:44:35	U	SMPL
ChkStd - 1	ChkStd	19041	6.4137	-	7470A	PPB	22 Mar 2023 17:46:58	U	SMPL
PB151609BL - 1	PBW	-87	-0.0598	-	7470A	PPB	22 Mar 2023 17:49:17	U	SMPL
PB151609BS - 1	LCSW	11078	3.7188	-	7470A	PPB	22 Mar 2023 17:51:34	U	SMPL
O1946-02 - 1	NWB-1895	131	0.0140	-	7470A	PPB	22 Mar 2023 17:53:51	U	SMPL
O1946-03 - 1	NWB-1896	566	0.1612	-	7470A	PPB	22 Mar 2023 17:56:08	U	SMPL
O2012-03 - 1	EB-2023-3-19	24	-0.0222	-	7470A	PPB	22 Mar 2023 18:01:20	U	SMPL
O2012-05 - 1	B-23-GW01	688	0.2025	-	7470A	PPB	22 Mar 2023 18:03:35	U	SMPL
CCV31 - 1	CCV31	16149	5.4350	-	7470A	PPB	22 Mar 2023 18:11:20	U	SMPL
CCB31 - 1	CCB31	-42	-0.0445	-	7470A	PPB	22 Mar 2023 18:13:35	U	SMPL
O2013-03 - 1	218	33	-0.0191	-	7470A	PPB	22 Mar 2023 18:15:52	U	SMPL
O2013-04 - 1	219	-18	-0.0364	-	7470A	PPB	22 Mar 2023 18:18:08	U	SMPL
O2013-08 DLX5	215-217	14579	4.9037	-	7470A	PPB	22 Mar 2023 18:20:24	U	SMPL
O2014-01 - 1	1711	63	-0.0090	-	7470A	PPB	22 Mar 2023 18:22:40	U	SMPL
O2020-41 - 1	FRAC-TANK-A821	149	0.0201	-	7470A	PPB	22 Mar 2023 18:24:59	U	SMPL
O2027-01 - 1	COMP-1	34	-0.0188	-	7470A	PPB	22 Mar 2023 18:27:15	U	SMPL
O2027-01DUP - 1	COMP-1DUP	61	-0.0097	-	7470A	PPB	22 Mar 2023 18:29:32	U	SMPL
O2027-01MS - 1	COMP-1MS	10004	3.3553	-	7470A	PPB	22 Mar 2023 18:31:49	U	SMPL
O2027-01MSD - 1	COMP-1MSD	10464	3.5110	-	7470A	PPB	22 Mar 2023 18:34:06	U	SMPL
O2027-03 - 1	COMP-2	-40	-0.0439	-	7470A	PPB	22 Mar 2023 18:36:23	U	SMPL
CCV32 - 1	CCV32	15233	5.1250	-	7470A	PPB	22 Mar 2023 18:41:22	U	SMPL
CCB32 - 1	CCB32	-71	-0.0543	-	7470A	PPB	22 Mar 2023 18:43:38	U	SMPL
O2027-05 - 1	COMP-3	38	-0.0175	-	7470A	PPB	22 Mar 2023 18:45:56	U	SMPL
O2027-01LX5 - 1	COMP-1L	-11	-0.0340	-	7470A	PPB	22 Mar 2023 18:48:12	U	SMPL
O2027-01A - 1	COMP-1A	12314	4.1371	-	7470A	PPB	22 Mar 2023 18:55:53	U	SMPL
CCV33 - 1	CCV33	15412	5.1856	-	7470A	PPB	22 Mar 2023 18:58:09	U	SMPL
CCB33 - 1	CCB33	-75	-0.0557	-	7470A	PPB	22 Mar 2023 19:00:26	U	SMPL