



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 3/23/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:50
In Date: 03/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 03/23/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124636

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1982-01	EYGB1	1	1.16	8.37	9.53	7.97	81.4	
O1982-02	EYGB2	2	1.17	8.63	9.8	8.19	81.3	
O1982-03	EYGB3	3	1.16	8.66	9.82	8.5	84.8	
O1982-04	EYGB4	4	1.16	8.64	9.8	8.53	85.3	
O1982-05	EYGB5	5	1.16	8.66	9.82	7.97	78.6	
O1982-06	EYGB6	6	1.16	8.77	9.93	7.99	77.9	
O1982-07	EYG98	7	1.15	8.79	9.94	8.82	87.3	
O1982-08	EYGB7	8	1.15	8.58	9.73	7.97	79.5	
O1982-09	EYGB8	9	1.16	8.82	9.98	8.54	83.7	
O1982-10	EYGB9	10	1.15	8.37	9.52	7.52	76.1	
O1982-11	EYGC0	11	1.15	8.62	9.77	8.23	82.1	
O1982-12	EYGC1	12	1.16	8.39	9.55	7.72	78.2	
O1982-13	EYGC2	13	1.15	8.57	9.72	7.87	78.4	
O1982-14	EYGC3	14	1.18	8.46	9.64	8.04	81.1	
O1982-15	EYGC4	15	1.16	8.75	9.91	8.33	81.9	
O1982-16	EYGC5	16	1.17	8.81	9.98	7.76	74.8	
O1982-17	EYGC6	17	1.15	8.48	9.63	7.77	78.1	
O1982-18	EYGC7	18	1.19	8.41	9.6	8.18	83.1	
O1982-19	EYGC8	19	1.16	8.74	9.9	7.92	77.3	
O1982-20	EYGC8MS	20	1.16	8.74	9.9	7.92	77.3	
O1982-21	EYGC8MSD	21	1.16	8.74	9.9	7.92	77.3	
O1982-22	EYGC9	22	1.16	8.81	9.97	8.42	82.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$