

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
CCV1	10.0189	0.0683	0.68	TOC	
CCB1	0.4087	0.1426	34.88	TOC	
LB124637BLW	0.3271	0.0960	29.34	TOC	
LB124637BSW.....	9.8105...	0.1958..	2.00...	TOC	..
01627-12	0.5158	0.1175	22.78	TOC	
01953-04	41.6297	0.6070	1.46	TOC	
01953-04DLX5.....	10.7873...	0.1929..	1.79...	TOC	..
02020-41	3.0196	0.3602	11.93	TOC	
02020-41MS	13.6267	0.4032	2.96	TOC	
02020-41MSD.....	13.7493...	0.2721..	1.98...	TOC	..
CCV2	10.4332	0.1148	1.10	TOC	
CCB2	0.4090	0.0690	16.86	TOC	

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	11	2023/03/23 09:25	
TOC 0 - 20 ppmC	Sample	12	2023/03/23 09:48	
TOC 0 - 20 ppmC	Sample	13	2023/03/23 10:10	
TOC 0 - 20 ppmC	...Sample	.. 14..	2023/03/23 10:43	..
TOC 0 - 20 ppmC	Sample	15	2023/03/23 11:43	
TOC 0 - 20 ppmC	Sample	16	2023/03/23 12:16	
TOC 0 - 20 ppmC	...Sample	.. 17..	2023/03/23 12:47	..
TOC 0 - 20 ppmC	Sample	19	2023/03/23 13:36	
TOC 0 - 20 ppmC	Sample	20	2023/03/23 14:07	
TOC 0 - 20 ppmC	...Sample	.. 20..	2023/03/23 14:37	..
TOC 0 - 20 ppmC	Sample	11	2023/03/23 15:07	
TOC 0 - 20 ppmC	Sample	12	2023/03/23 15:29	

```
=====
Sample ID:   CCV1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   03230830
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/03/23 09:25
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	9.9343	4.9672	431909	0.509	0.708	245
2	10.0401	5.0200	436505	0.429	0.628	255
3	10.0037	5.0018	434923	0.361	0.561	233
4	10.0974	5.0487	438997	0.308	0.508	252

```
-----
<<<Statistics>>>   Mean:  10.0189   Std Dev:  0.0683   RSD: 0.68
=====
```

```
=====
Sample ID:   CCB1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   03230830
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/03/23 09:48
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.6113	0.3057	26577	0.107	0.307	113
2	0.3898	0.1949	16949	0.148	0.348	107
3	0.3537	0.1769	15379	0.137	0.334	106
4	0.2801	0.1400	12177	0.156	0.355	105

```
-----
<<<Statistics>>>   Mean:   0.4087   Std Dev:  0.1426   RSD: 34.88
=====
```

```
=====
Sample ID:   LB124637BLW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   03230830
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/03/23 10:10
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.4659	0.2329	20255	-0.134	0.065	111
2	0.2447	0.1223	10638	-0.083	0.117	105
3	0.3010	0.1505	13084	-0.137	0.062	106
4	0.2970	0.1485	12914	-0.173	0.024	107

```
-----
<<<Statistics>>>   Mean:   0.3271   Std Dev:  0.0960   RSD: 29.34
=====
```

```
=====
Sample ID:   LB124637BSW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   03230830
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/03/23 10:43
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	9.6197	4.8098	418229	-0.157	0.041	235
2	9.6735	4.8368	420570	-0.196	0.004	248
3	9.9193	4.9596	431254	-0.246	-0.048	260
4	10.0296	5.0148	436049	-0.301	-0.102	258

```
-----
<<<Statistics>>>   Mean:   9.8105   Std Dev:  0.1958   RSD: 2.00
=====
```

```
=====
Sample ID:   01627-12                 Mode:      TOC
=====
```

Sample ID:	02020-41MS	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	03231121
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2023/03/23 14:07
Operator ID:	IZ MA	Sample Type:	Sample

```
<<<Statistics>>>      Mean:   13.6267      Std Dev:   0.4032      RSD: 2.96
```

```
Mode:          TOC
Filename:      03231121
Timestamp:     2023/03/23 14:37
Sample Type:   Sample
```

```
<<<Statistics>>>    Mean:   13.7493    Std Dev:   0.2721    RSD: 1.98
```

```
Mode:          TOC
Filename:      03231121
Timestamp:     2023/03/23 15:07
```

```
<<<Statistics>>>    Mean:    10.4332    Std Dev:    0.1148    RSD: 1.10
```

```
Mode:          TOC
Filename:      03231121
Timestamp:     2023/03/23 15:29
Sample Type:   Sample
```

```
<<<Statistics>>>    Mean:    0.4090    Std Dev:    0.0690    RSD: 16.86
```

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	12656	4375	34.57	TOC	
0.5PPM	25627	6271	24.47	TOC	
1.0PPM	48950	3630	7.42	TOC	
2.0PPM.....	94452...	5974..	6.32...	TOC	..
5.0PPM	235780	3567	1.51	TOC	
10.0PPM	450480	4608	1.02	TOC	
20.0PPM.....	874934...	15444..	1.77...	TOC	..
ICV1	10.4757	0.1465	1.40	TOC	
ICB1	0.3431	0.0262	7.64	TOC	
IC-20.....	0.2487...	0.0423..	16.99...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2023/01/03 13:02	
TOC 0 - 20 ppmC	TOC Standard	2	2023/01/03 13:25	
TOC 0 - 20 ppmC	TOC Standard	3	2023/01/03 14:39	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2023/01/03 15:03	..
TOC 0 - 20 ppmC	TOC Standard	5	2023/01/03 15:31	
TOC 0 - 20 ppmC	TOC Standard	6	2023/01/03 15:59	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2023/01/03 16:28	..
TOC 0 - 20 ppmC	Sample	8	2023/01/03 16:57	
TOC 0 - 20 ppmC	Sample	9	2023/01/03 17:19	
TOC 0 - 20 ppmC	...Sample	.. 10..	2023/01/03 17:42	..

```
=====
Sample ID:   0.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   01031241
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/01/03 13:02
Operator ID: IZ AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			12151	0.978	1.178	104
2			10509	0.984	1.182	102
3			18935	0.958	1.157	102
4			9031	0.913	1.109	103

```
-----
<<<Statistics>>>   Mean:    12656   Std Dev:    4375   RSD: 34.57
=====
```

```
=====
Sample ID:   0.5PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   01031241
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/01/03 13:25
Operator ID: IZ AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			34726	0.734	0.930	118
2			20679	0.795	0.993	112
3			24570	0.784	0.981	114
4			22532	0.762	0.961	114

```
-----
<<<Statistics>>>   Mean:    25627   Std Dev:    6271   RSD: 24.47
=====
```

```
=====
Sample ID:   1.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   01031241
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/01/03 14:39
Operator ID: IZ AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			52000	0.472	0.671	128
2			44159	0.544	0.743	118
3			48121	0.501	0.699	120
4			51518	0.471	0.671	121

```
-----
<<<Statistics>>>   Mean:    48950   Std Dev:    3630   RSD: 7.42
=====
```

```
=====
Sample ID:   2.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   01031241
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/01/03 15:03
Operator ID: IZ AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			101603	0.389	0.589	144
2			94439	0.408	0.608	125
3			94785	0.393	0.592	135
4			86981	0.421	0.617	124

```
-----
<<<Statistics>>>   Mean:    94452   Std Dev:    5974   RSD: 6.32
=====
```

```
=====
Sample ID:   5.0PPM                      Mode:      TOC
=====
```

Sample ID:	ICB1	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	01031241
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2023/01/03 17:19
Operator ID:	IZ AP	Sample Type:	Sample

```
<<<Statistics>>>      Mean:      0.3431      Std Dev:      0.0262      RSD: 7.64
```

```
Mode:          TOC
Filename:      01031241
Timestamp:     2023/01/03 17:42
Sample Type:   Sample
```

```
<<<Statistics>>>      Mean:      0.2487      Std Dev:      0.0423      RSD: 16.99
```

Cal. Curve ID: TOC WATER 0-20PPM
 Created: 2023/01/03 16:29
 Calibration Factor (m): 8.695e+04
 Y Intercept (b): 9863
 r-squared: 0.99968

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message <i>re</i>	Date & Time
0.0PPM	12656	0.000	0.032	—	2023/01/03 13:02
0.5PPM	25627	0.250	0.181	-27.6	2023/01/03 13:25
1.0PPM	48949	0.500	0.450	-10.0	2023/01/03 14:39
2.0PPM	94452	1.000	0.973	-2.7	2023/01/03 15:03
5.0PPM	235780	2.500	2.598	3.9	2023/01/03 15:31
10.0PPM	450480	5.000	5.067	1.3	2023/01/03 15:59
20.0PPM	874934	10.000	9.949	-0.5	2023/01/03 16:28

12

1/3/23