



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB124638
Analysis Date: 03/23/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: Iwona
Extraction Date: 03/23/2023
Extraction IN Time: 08:25
Extraction OUT Time: 08:45
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB124638BL	LB124638BL	WATER	1.3	1000	100	2.6587	2.6587	3.02	2.6588	2.6588	0.0001	0.1
2	LB124638BS	LB124638BS	WATER	1.3	1000	100	2.7896	2.7896	3.01	2.8066	2.8066	0.0170	17
3	LB124638BSD	LB124638BSD	WATER	1.3	1000	100	3.0258	3.0258	3.03	3.0431	3.0431	0.0173	17.3
4	O2020-41	FRAC-TANK-A821	WATER	1.3	1000	100	2.9820	2.9820	3.04	2.9844	2.9844	0.0024	2.4
5	O2040-01	GRAB	WATER	1.6	1000	100	3.0322	3.0322	3.05	3.0987	3.0987	0.0665	66.5

QC Batch# LB124638Test: TPHAnalysis Date: 03/23/2023**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2312
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator DataAnalytical Balance ID # : WC SC-6Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:26
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time1: 09:30
Bal Check Time: 08:35 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:00
Out Time1: 10:25

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:01
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time2: 11:30
Bal Check Time: 12:45 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 12:38
Out Time2: 12:00