



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 3/23/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 15:30

In Date: 03/22/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:38

Out Date: 03/23/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB124639

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1992-01	EYG98	1	1.15	8.79	9.94	8.82	87.3	
O1992-02	EYGA8	2	1.15	8.65	9.8	8.55	85.5	
O1992-04	EYGB0	3	1.15	8.58	9.73	7.96	79.4	
O1992-05	EYGB1	4	1.16	8.37	9.53	7.97	81.4	
O1992-06	EYGB2	5	1.17	8.63	9.8	8.19	81.3	
O1992-07	EYGB3	6	1.16	8.66	9.82	8.5	84.8	
O1992-08	EYGB4	7	1.16	8.64	9.8	8.53	85.3	
O1992-09	EYGB5	8	1.16	8.66	9.82	7.97	78.6	
O1992-10	EYGB6	9	1.16	8.77	9.93	7.99	77.9	
O1992-11	EYGB7	10	1.15	8.58	9.73	7.97	79.5	
O1992-12	EYGB8	11	1.16	8.82	9.98	8.54	83.7	
O1992-13	EYGB9	12	1.15	8.37	9.52	7.52	76.1	
O1992-14	EYGC0	13	1.15	8.62	9.77	8.23	82.1	
O1992-15	EYGC1	14	1.16	8.39	9.55	7.72	78.2	
O1992-16	EYGC2	15	1.15	8.57	9.72	7.87	78.4	
O1992-17	EYGC3	16	1.18	8.46	9.64	8.04	81.1	
O1992-18	EYGC4	17	1.16	8.75	9.91	8.33	81.9	
O1992-19	EYGC5	18	1.17	8.81	9.98	7.76	74.8	
O1992-20	EYGC8	19	1.16	8.74	9.9	7.92	77.3	
O1992-21	EYGC8MS	20	1.16	8.74	9.9	7.92	77.3	
O1992-22	EYGC8MSD	21	1.16	8.74	9.9	7.92	77.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$