



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124647

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	03/23/2023	13:45
2	O2020-41		1	14	100.00	>212.0	>212.0	03/23/2023	14:15
3	O2020-41DUP		1	14	100.00	>212.0	>212.0	03/23/2023	14:45
4	O2027-01		1	12	100.00	>212.0	>212.0	03/23/2023	15:15
5	O2027-03		1	13	100.00	>212.0	>212.0	03/23/2023	15:45
6	O2027-05		1	14	100.00	>212.0	>212.0	03/23/2023	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-3-22

WorkList ID : 168381

Department : Wet-Chemistry

Date : 03-22-2023 08:24:58

16124647

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/24/2023	Water	O2020-41	Flash Point	Cool 4 deg C	PSEG03	K11	FRAC-TANK-A821	03/20/2023	1010B
03/28/2023	Water	O2027-01	Flash Point	Cool 4 deg C	PSEG03	L11	COMP-1	03/21/2023	1010B
03/28/2023	Water	O2027-03	Flash Point	Cool 4 deg C	PSEG03	L11	COMP-2	03/21/2023	1010B
03/28/2023	Water	O2027-05	Flash Point	Cool 4 deg C	PSEG03	L11	COMP-3	03/21/2023	1010B

Date/Time 03/23/2023 13:30
 Raw Sample Received by: RMA (w/c)
 Raw Sample Relinquished by: RMA (w/c)

Date/Time 03/23/2023 16:50
 Raw Sample Received by: RMA (w/c)
 Raw Sample Relinquished by: RMA (w/c)