

**PERCENT SOLID****Supervisor:** Iwona**Analyst:** JIGNESH**Date:** 3/24/2023**OVENTEMP IN Celsius(°C):** 107**Time IN:** 17:00**In Date:** 03/23/2023**Weight Check 1.0g:** 1.00**Weight Check 10g:** 10.00**OvenID:** M OVEN-1**OVENTEMP OUT Celsius(°C):** 103**Time OUT:** 08:05**Out Date:** 03/24/2023**Weight Check 1.0g:** 1.00**Weight Check 10g:** 10.00**BalanceID:** M SC-4**Thermometer ID:** %SOLIDS-OVEN**QC:**LB124648

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2047-01	R-PIT	1	1.14	8.60	9.74	8.1	80.9	
O2049-01	OK-02-032323	3	1.12	8.70	9.82	8.79	88.2	
O2049-02	OK-02-032323-EPH-2	4	1.12	8.70	9.82	8.79	88.2	
O2050-01	KMH349G-1-1	12	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-02	KMH349G-1-2	13	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-03	BC260963-1-1	14	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-04	BC260963-1-2	15	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-05	BC260963-2-1	16	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-06	BC260963-2-2	17	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-07	ECH036X-1-1	18	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-08	ECH036X-1-2	19	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-09	JMA187K-1-1	20	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-10	JMA187K-1-2	21	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-11	JMA187K-2-1	22	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2050-12	JMA187K-2-2	23	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2051-01	BC226716-1-1	24	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2051-02	BC226716-1-2	25	1.00	1.00	2.00	2.00	100.0	PILC REEL
O2058-01	CHAMBER-B-72-EAST	2	1.15	8.80	9.95	5.33	47.5	SLUDGE SAMPLE
O2060-01	31	5	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2060-02	32	6	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2060-03	33	7	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2060-04	34	8	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2060-05	35	9	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2060-06	36	10	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O2068-01	CONCRETE-TCLP	11	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O2070-01	SOILPILE-1-0323	26	1.18	8.57	9.75	8.26	82.6	
O2070-03	SOILPILE-2-0323	27	1.13	8.84	9.97	9.17	91.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$