


7471B

LB124650 INSTRUMENT: CV1.

Linear

 μ Abs.:

39512



A= 0.0000e+000

B= 2.4848e-004

C= 3.9771e-002

Rho= 0.9976911

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.056	0.056	64	0.000	64					
0.2	0.200	0.213	0.013	697	0.0 %	697					
2.5	2.500	2.619	0.119	10379	0.0 %	10379					
5.0	5.000	4.566	-0.434	18216	0.0 %	18216					
7.5	7.500	7.889	0.389	31587	0.0 %	31587					
10.0	10.000	9.858	-0.142	39512	0.0 %	39512					

7
5
9
+55
-2

3/23/23

LB124650 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Mar 2023 20:08:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	64	-	0.0000	7471B	PPB	23 Mar 2023 20:17:10	S	Std
0.2 - 1	50.2	697	-	0.2000	7471B	PPB	23 Mar 2023 20:19:27	S	Std
2.5 - 1	52.5	10379	-	2.5000	7471B	PPB	23 Mar 2023 20:21:45	S	Std
5.0 - 1	55.0	18216	-	5.0000	7471B	PPB	23 Mar 2023 20:24:02	S	Std
7.5 - 1	57.5	31587	-	7.5000	7471B	PPB	23 Mar 2023 20:26:19	S	Std
10.0 - 1	510.0	39512	-	10.0000	7471B	PPB	23 Mar 2023 20:45:17	S	Std
ICV02 - 1	ICV02	16799	4.2140	-	7471B	PPB	23 Mar 2023 20:51:53	U	SMPL
ICB02 - 1	ICB02	-53	0.0266	-	7471B	PPB	23 Mar 2023 20:54:08	U	SMPL
CCV06 - 1	CCV06	21761	5.4470	-	7471B	PPB	23 Mar 2023 21:00:06	U	SMPL
CCB06 - 1	CCB06	-66	0.0234	-	7471B	PPB	23 Mar 2023 21:02:21	U	SMPL
CRA - 1	CRA	559	0.1787	-	7471B	PPB	23 Mar 2023 21:04:38	U	SMPL
HighStd - 1	HighStd	31897	7.9656	-	7471B	PPB	23 Mar 2023 21:06:53	U	SMPL
ChkStd - 1	ChkStd	24648	6.1644	-	7471B	PPB	23 Mar 2023 21:09:09	U	SMPL
PB151626BL - 1	PB151626BL	-111	0.0122	-	7471B	PPB	23 Mar 2023 21:11:27	U	SMPL
PB151626BS - 1	PB151626BS	14564	3.6587	-	7471B	PPB	23 Mar 2023 21:13:45	U	SMPL
O2022-01 - 1	AU-05-32123	825	0.2448	-	7471B	PPB	23 Mar 2023 21:16:02	U	SMPL
O2022-01DUP - 1	AU-05-32123DUP	631	0.1966	-	7471B	PPB	23 Mar 2023 21:18:20	U	SMPL
O2022-01MS - 1	AU-05-32123MS	17700	4.4379	-	7471B	PPB	23 Mar 2023 21:28:43	U	SMPL
O2022-01MSD - 1	AU-05-32123MSD	17129	4.2960	-	7471B	PPB	23 Mar 2023 21:31:01	U	SMPL
O2028-01 - 1	TR-05-032123	2766	0.7271	-	7471B	PPB	23 Mar 2023 21:33:19	U	SMPL
CCV07 - 1	CCV07	18597	4.6608	-	7471B	PPB	23 Mar 2023 21:43:53	U	SMPL
CCB07 - 1	CCB07	-69	0.0226	-	7471B	PPB	23 Mar 2023 21:46:08	U	SMPL
O2029-01 - 1	OR-2-032123	1005	0.2895	-	7471B	PPB	23 Mar 2023 21:48:26	U	SMPL
O2032-02 - 1	S-2	1624014	HIGH	-	7471B	PPB	23 Mar 2023 21:50:42	X	
O2045-06 - 1	WC1-20230322	12627	3.1774	-	7471B	PPB	23 Mar 2023 21:58:43	U	SMPL
O2045-08 - 1	WC2-20230322	2762	0.7261	-	7471B	PPB	23 Mar 2023 22:01:04	U	SMPL
O2032-02DLX100 - 1		183082	45.5326	-	7471B	PPB	23 Mar 2023 22:05:25	U	SMPL
CCV08 - 1	CCV08	18863	4.7269	-	7471B	PPB	23 Mar 2023 22:23:41	U	SMPL
CCB08 - 1	CCB08	-255	-0.0236	-	7471B	PPB	23 Mar 2023 22:25:58	U	SMPL
O2022-01LX5 - 1		-90	0.0174	-	7471B	PPB	23 Mar 2023 22:28:15	U	SMPL
O2033-01 - 1		1561	0.4277	-	7471B	PPB	23 Mar 2023 22:30:31	U	SMPL
O2036-01 - 1		18	0.0442	-	7471B	PPB	23 Mar 2023 22:32:48	U	SMPL
O2032-02DL2X500 - 1		34603	8.6380	-	7471B	PPB	23 Mar 2023 22:38:28	U	SMPL
CCV09 - 1	CCV09	18924	4.7421	-	7471B	PPB	23 Mar 2023 22:49:03	U	SMPL
CCB09 - 1	CCB09	-445	-0.0708	-	7471B	PPB	23 Mar 2023 22:54:53	U	SMPL