



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124664

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/24/2023	06:20
2	CAL2	1	Water	NA	NA	20.6	7.01	03/24/2023	06:21
3	CAL3	1	Water	NA	NA	20.7	10.02	03/24/2023	06:22
4	ICV	1	Water	NA	NA	20.6	6.99	03/23/2023	06:25
5	CCV1	1	Water	NA	NA	20.7	2.01	03/24/2023	06:30
6	O2056-01	1	Water	NA	NA	20.7	7.64	03/24/2023	06:34
7	O2056-01DUP	1	Water	NA	NA	20.6	7.65	03/24/2023	06:35
8	O2069-01	1	Water	NA	NA	20.9	7.35	03/24/2023	06:37
9	O2069-02	1	Water	NA	NA	20.7	7.84	03/24/2023	06:38
10	CCV2	1	Water	NA	NA	20.8	12.02	03/24/2023	06:40

VB 124664

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2056 w WorkList ID : 168448 Department : Wet-Chemistry Date : 03-24-2023 05:54:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/22/2023	Water	O2056-01	pH	Cool 4 deg C	MINO01	M12	025-OUTFALL	03/22/2023	9040C
03/22/2023	Water	O2069-01	pH	Cool 4 deg C	ENV170	M11	API-3-3	03/22/2023	9040C
03/22/2023	Water	O2069-02	pH	Cool 4 deg C	ENV170	M11	API-3-4	03/22/2023	9040C

Date/Time 03/24/23 06:10

Raw Sample Received by: JH (CEN)

Raw Sample Relinquished by: JH (CEN)

Date/Time 03/24/23 09:10

Raw Sample Received by: JH (CEN)

Raw Sample Relinquished by: JH (CEN)