



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124670

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/24/2023	09:30
2	O2052-03		1	12	100.00	>212.0	>212.0	03/24/2023	10:00
3	O2052-04		1	14	100.00	>212.0	>212.0	03/24/2023	10:30
4	O2065-03		1	12	100.00	>212.0	>212.0	03/24/2023	11:00
5	O2073-03		1	13	100.00	>212.0	>212.0	03/24/2023	11:30
6	O2073-03DUP		1	14	100.00	>212.0	>212.0	03/24/2023	12:00
7	O2073-04		1	14	100.00	>212.0	>212.0	03/24/2023	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124670

WorkList Name : fp3-24

WorkList ID : 168450

Department : Wet-Chemistry

Date : 03-24-2023 08:24:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/28/2023	Water	O2052-03	Flash Point	Cool 4 deg C	PSEG03	M12	AUD-1699	03/23/2023	1010B
03/28/2023	Water	O2052-04	Flash Point	Cool 4 deg C	PSEG03	M12	AUD-1700	03/23/2023	1010B
03/28/2023	Water	O2065-03	Flash Point	Cool 4 deg C	PSEG03	M11	1545	03/23/2023	1010B
03/28/2023	Water	O2073-03	Flash Point	Cool 4 deg C	PSEG03	M11	TRE-1717	03/23/2023	1010B
03/28/2023	Water	O2073-04	Flash Point	Cool 4 deg C	PSEG03	M11	TRE-1660	03/23/2023	1010B

Date/Time 03/23/2023 9:15
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: JH CWG

Date/Time 03/24/2023 13:10
 Raw Sample Received by: JH CWG
 Raw Sample Relinquished by: RM CWG