



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124676

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP101696
Turbidity Calibration std, 20NTU	WP101691
Turbidity Calibration std, 40NTU	WP101690
Turbidity Calibration std, 1NTU	WP101694
Turbidity Calibration std, 5NTU	WP101693
Turbidity Calibration - CCV std, 10 NTU	WP101692
10 NTU Standard 500 ml	W2971

Intercept: 0.023

Slope: 0.9303

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.152	0.14		03/24/2023	16:00
2	CAL2	1	1	1.150	1.21	21.1	03/24/2023	16:03
3	CAL3	5	1	4.390	4.69	-6.1	03/24/2023	16:07
4	CAL4	10	1	9.450	10.13	1.3	03/24/2023	16:10
5	CAL5	20	1	18.300	19.65	-1.8	03/24/2023	16:13
6	CAL6	40	1	37.400	40.18	0.4	03/24/2023	16:17



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124676

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.310	9.983	03/24/2023	16:20
2	ICB		1	0.142	0.128	03/24/2023	16:23
3	CCV1	10	1	9.140	9.800	03/24/2023	16:27
4	CCB1		1	0.148	0.134	03/24/2023	16:30
5	LB124676BL		1	0.156	0.143	03/24/2023	16:33
6	O2096-01		1	0.509	0.522	03/24/2023	16:37
7	O2096-01DUP		1	0.511	0.525	03/24/2023	16:40
8	O2096-03		1	0.344	0.345	03/24/2023	16:43
9	O2096-05		1	0.339	0.340	03/24/2023	16:47
10	O2096-07		1	5.050	5.404	03/24/2023	16:50
11	CCV2	10	1	9.230	9.897	03/24/2023	16:53
12	CCB2		1	0.150	0.137	03/24/2023	16:57

WORKLIST(Hardcopy Internal Chain)

LB124676

WorkList Name : TURB-032423

WorkList ID : 168489

Department : Wet-Chemistry

Date : 03-24-2023 15:49:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/25/2023	Water	O2096-01	Turbidity	Cool 4 deg C	LOCK01	M12	MW-1	03/23/2023	SM2130 B
03/25/2023	Water	O2096-03	Turbidity	Cool 4 deg C	LOCK01	M12	MW-3	03/23/2023	SM2130 B
03/25/2023	Water	O2096-05	Turbidity	Cool 4 deg C	LOCK01	M12	MW-4	03/23/2023	SM2130 B
03/25/2023	Water	O2096-07	Turbidity	Cool 4 deg C	LOCK01	M12	MW-2	03/23/2023	SM2130 B

Date/Time 03/24/23 15:50
 Raw Sample Received by: 12/507
 Raw Sample Relinquished by: 12/507

Date/Time 03/24/23 17:10
 Raw Sample Received by: 12/507
 Raw Sample Relinquished by: 12/507