



Analytical Summary Report

Analysis Method: SM 4500-H B

Parameter: pH

Run Number: LB124679

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/24/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	03/24/2023	16:11
3	CAL3	1	Water	NA	NA	20.7	10.03	03/24/2023	16:13
4	ICV	1	Water	NA	NA	20.7	6.98	03/24/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	03/24/2023	16:20
6	O2096-01	1	Water	NA	NA	20.1	6.06	03/24/2023	16:25
7	O2096-03	1	Water	NA	NA	20.5	6.87	03/24/2023	16:30
8	O2096-05	1	Water	NA	NA	20.6	6.89	03/24/2023	16:33
9	O2096-07	1	Water	NA	NA	20.7	6.98	03/24/2023	16:39
10	O2096-07DUP	1	Water	NA	NA	20.8	6.99	03/24/2023	16:40
11	CCV2	1	Water	NA	NA	20.6	12.03	03/24/2023	16:44

WORKLIST(Hardcopy Internal Chain)

VB124679

WorkList Name : ph o2096

WorkList ID : 168490

Department : Wet-Chemistry

Date : 03-24-2023 15:49:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/23/2023	Water	O2096-01	pH	✓	Cool 4 deg C	LOCK01	M12	MW-1	03/23/2023 SM 4500-H B
03/23/2023	Water	O2096-03	pH	✓	Cool 4 deg C	LOCK01	M12	MW-3	03/23/2023 SM 4500-H B
03/23/2023	Water	O2096-05	pH	✓	Cool 4 deg C	LOCK01	M12	MW-4	03/23/2023 SM 4500-H B
03/23/2023	Water	O2096-07	pH	✓	Cool 4 deg C	LOCK01	M12	MW-2	03/23/2023 SM 4500-H B

Date/Time 03/24/23 15:55
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM

Date/Time 03/24/23 17:30
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM