


LB124686 Instrument: CV1

SFAM01.1

Linear



A= 0.0000e+000

B= 1.6412e-004

C= -3.3173e-002

Rho= 0.9996534

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.020	0.020	325	0.000	325				
0.05	0.050					185	0			
0.20	0.200	0.167	-0.033	1222	0.0 %	1222				
2.50	2.500	2.568	0.068	15849	0.0 %	15849				
5.00	5.000	4.854	-0.146	29780	0.0 %	29780				
7.50	7.500	7.657	0.157	46860	0.0 %	46860				
10.0	10.000	9.933	-0.067	60724	0.0 %	60724				

LB124686 INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0	50.0	325	-		0 SFAM01.1	PPB	3/26/2023 19:35	Std
0.2	50.2	1222	-		0.2 SFAM01.1	PPB	3/26/2023 19:38	Std
2.5	52.5	15849	-		2.5 SFAM01.1	PPB	3/26/2023 19:40	Std
5	55.0	29780	-		5 SFAM01.1	PPB	3/26/2023 19:42	Std
7.5	57.5	46860	-		7.5 SFAM01.1	PPB	3/26/2023 19:44	Std
10	510.0	60724	-		10 SFAM01.1	PPB	3/26/2023 19:47	Std
ICV001	ICV001	25597	4.1678 -		SFAM01.1	PPB	3/26/2023 19:52	SMPL
ICB001	ICB001	357	0.0254 -		SFAM01.1	PPB	3/26/2023 19:54	SMPL
CCV001	CCV001	31949	5.2103 -		SFAM01.1	PPB	3/26/2023 19:56	SMPL
CCB001	CCB001	289	0.0143 -		SFAM01.1	PPB	3/26/2023 19:58	SMPL
PB151709BL	PBW	502	0.0492 -		SFAM01.1	PPB	3/26/2023 20:01	SMPL
O1914-01	YBL54	13286	2.1473 -		SFAM01.1	PPB	3/26/2023 20:03	SMPL
O1914-02	YBL56	35720	5.8292 -		SFAM01.1	PPB	3/26/2023 20:05	SMPL
O1914-03	YBL58	37131	6.0607 -		SFAM01.1	PPB	3/26/2023 20:07	SMPL
O1914-04	YBL64	34844	5.6854 -		SFAM01.1	PPB	3/26/2023 20:10	SMPL
O1914-05	YBL78	16744	2.7148 -		SFAM01.1	PPB	3/26/2023 20:12	SMPL
O1914-06	YBL80	6775	1.0787 -		SFAM01.1	PPB	3/26/2023 20:14	SMPL
O1914-07	YBL82	16864	2.7345 -		SFAM01.1	PPB	3/26/2023 20:17	SMPL
O1914-08	YBL84	8583	1.3755 -		SFAM01.1	PPB	3/26/2023 20:19	SMPL
O1914-09	YBL88	4353	0.6812 -		SFAM01.1	PPB	3/26/2023 20:21	SMPL
O1914-10	YBL88D	3467	0.5358 -		SFAM01.1	PPB	3/26/2023 20:23	SMPL
O1914-11	YBL88S	8803	1.4116 -		SFAM01.1	PPB	3/26/2023 20:26	SMPL
O1914-12	YBL92	4540	0.7119 -		SFAM01.1	PPB	3/26/2023 20:28	SMPL
O1914-13	YBL94	4538	0.7116 -		SFAM01.1	PPB	3/26/2023 20:30	SMPL
O1914-14	YBL66	21649	3.5198 -		SFAM01.1	PPB	3/26/2023 20:32	SMPL
O1914-15	YBL68	27908	4.5471 -		SFAM01.1	PPB	3/26/2023 20:35	SMPL
O1914-16	YBL74	17841	2.8949 -		SFAM01.1	PPB	3/26/2023 20:37	SMPL
O1914-17	YBL86	2732	0.4152 -		SFAM01.1	PPB	3/26/2023 20:39	SMPL
O1914-18	YBL98	9686	1.5565 -		SFAM01.1	PPB	3/26/2023 20:42	SMPL
O1914-19	YBLA0	7923	1.2671 -		SFAM01.1	PPB	3/26/2023 20:44	SMPL
CCV002	CCV002	28930	4.7148 -		SFAM01.1	PPB	3/26/2023 20:46	SMPL
CCB002	CCB002	244	0.0069 -		SFAM01.1	PPB	3/26/2023 20:48	SMPL
O1914-20	YBLA6	10083	1.6216 -		SFAM01.1	PPB	3/26/2023 20:51	SMPL
O1914-21	YBLA8	9969	1.6029 -		SFAM01.1	PPB	3/26/2023 20:53	SMPL
O1914-22	YBLB0	3296	0.5078 -		SFAM01.1	PPB	3/26/2023 20:55	SMPL
PB151710BL	PBW	453	0.0412 -		SFAM01.1	PPB	3/26/2023 20:57	SMPL
O1916-01	YBL55	1620	0.2327 -		SFAM01.1	PPB	3/26/2023 21:00	SMPL
O1916-02	YBL57	5565	0.8801 -		SFAM01.1	PPB	3/26/2023 21:02	SMPL
O1916-03	YBL59	2041	0.3018 -		SFAM01.1	PPB	3/26/2023 21:04	SMPL
O1916-04	YBL65	1698	0.2455 -		SFAM01.1	PPB	3/26/2023 21:06	SMPL
O1916-05	YBL79	739	0.0881 -		SFAM01.1	PPB	3/26/2023 21:09	SMPL
O1916-06	YBL81	691	0.0802 -		SFAM01.1	PPB	3/26/2023 21:11	SMPL
O1916-07	YBL83	702	0.082 -		SFAM01.1	PPB	3/26/2023 21:13	SMPL
O1916-08	YBL85	614	0.0676 -		SFAM01.1	PPB	3/26/2023 21:16	SMPL
O1916-09	YBL89	501	0.0491 -		SFAM01.1	PPB	3/26/2023 21:18	SMPL
O1916-10	YBL89D	472	0.0443 -		SFAM01.1	PPB	3/26/2023 21:20	SMPL

O1916-11	YBL89S	5581	0.8828 -	SFAM01.1	PPB	3/26/2023 21:22	SMPL
O1916-12	YBL93	511	0.0507 -	SFAM01.1	PPB	3/26/2023 21:25	SMPL
O1916-13	YBL95	544	0.0561 -	SFAM01.1	PPB	3/26/2023 21:27	SMPL
O1916-14	YBL67	630	0.0702 -	SFAM01.1	PPB	3/26/2023 21:29	SMPL
O1916-15	YBL69	1026	0.1352 -	SFAM01.1	PPB	3/26/2023 21:31	SMPL
O1916-16	YBL75	666	0.0761 -	SFAM01.1	PPB	3/26/2023 21:34	SMPL
CCV003	CCV003	30493	4.9713 -	SFAM01.1	PPB	3/26/2023 21:36	SMPL
CCB003	CCB003	309	0.0175 -	SFAM01.1	PPB	3/26/2023 21:38	SMPL
O1916-17	YBL87	618	0.0683 -	SFAM01.1	PPB	3/26/2023 21:41	SMPL
O1916-18	YBL99	682	0.0788 -	SFAM01.1	PPB	3/26/2023 21:43	SMPL
O1916-19	YBLA1	1029	0.1357 -	SFAM01.1	PPB	3/26/2023 21:45	SMPL
O1916-20	YBLA7	565	0.0596 -	SFAM01.1	PPB	3/26/2023 21:47	SMPL
O1916-21	YBLA9	479	0.0454 -	SFAM01.1	PPB	3/26/2023 21:50	SMPL
O1916-22	YBLB1	500	0.0489 -	SFAM01.1	PPB	3/26/2023 21:52	SMPL
PB151711BL	PBW	453	0.0412 -	SFAM01.1	PPB	3/26/2023 21:54	SMPL
O1915-01	YBLA2	12531	2.0234 -	SFAM01.1	PPB	3/26/2023 21:56	SMPL
O1915-02	YBLA2D	14929	2.417 -	SFAM01.1	PPB	3/26/2023 21:59	SMPL
O1915-03	YBLA2S	18664	3.0299 -	SFAM01.1	PPB	3/26/2023 22:01	SMPL
O1915-04	YBLB2	2408	0.362 -	SFAM01.1	PPB	3/26/2023 22:03	SMPL
O1915-05	YBLB4	6471	1.0288 -	SFAM01.1	PPB	3/26/2023 22:06	SMPL
O1915-06	YBLB6	4554	0.7142 -	SFAM01.1	PPB	3/26/2023 22:08	SMPL
O1917-01	YBLA3	278	0.0125 -	SFAM01.1	PPB	3/26/2023 22:10	SMPL
O1917-02	YBLA3D	332	0.0213 -	SFAM01.1	PPB	3/26/2023 22:12	SMPL
O1917-03	YBLA3S	5363	0.847 -	SFAM01.1	PPB	3/26/2023 22:15	SMPL
O1917-04	YBLB3	265	0.0103 -	SFAM01.1	PPB	3/26/2023 22:17	SMPL
O1917-05	YBLB5	384	0.0298 -	SFAM01.1	PPB	3/26/2023 22:19	SMPL
O1917-06	YBLB7	298	0.0157 -	SFAM01.1	PPB	3/26/2023 22:21	SMPL
O1917-07	YBLB8	261	0.0097 -	SFAM01.1	PPB	3/26/2023 22:24	SMPL
CCV004	CCV004	33249	5.4236 -	SFAM01.1	PPB	3/26/2023 22:26	SMPL
CCB004	CCB004	278	0.0125 -	SFAM01.1	PPB	3/26/2023 22:28	SMPL
O2062-01	MCB961	243	0.0067 -	SFAM01.1	PPB	3/26/2023 22:31	SMPL
O2062-02	MCB961D	255	0.0087 -	SFAM01.1	PPB	3/26/2023 22:33	SMPL
O2062-03	MCB961S	5215	0.8227 -	SFAM01.1	PPB	3/26/2023 22:35	SMPL
O2063-01	MCB962	162	-0.0066 -	SFAM01.1	PPB	3/26/2023 22:37	SMPL
O2063-02	MCB962D	210	0.0013 -	SFAM01.1	PPB	3/26/2023 22:40	SMPL
O2063-03	MCB962S	5381	0.85 -	SFAM01.1	PPB	3/26/2023 22:42	SMPL
O2088-01	YBR55	511	0.0507 -	SFAM01.1	PPB	3/26/2023 22:44	SMPL
O2088-02	YBR55D	620	0.0686 -	SFAM01.1	PPB	3/26/2023 22:47	SMPL
O2088-03	YBR55S	6378	1.0136 -	SFAM01.1	PPB	3/26/2023 22:49	SMPL
O2088-04	YBR59	222	0.0033 -	SFAM01.1	PPB	3/26/2023 22:51	SMPL
O2089-01	YBR56	294	0.0151 -	SFAM01.1	PPB	3/26/2023 22:53	SMPL
O2089-02	YBR56D	272	0.0115 -	SFAM01.1	PPB	3/26/2023 22:56	SMPL
O2089-03	YBR56S	5454	0.8619 -	SFAM01.1	PPB	3/26/2023 22:58	SMPL
O2089-04	YBR60	186	-0.0026 -	SFAM01.1	PPB	3/26/2023 23:00	SMPL
PB151712BL	PBW	411	0.0343 -	SFAM01.1	PPB	3/26/2023 23:02	SMPL
O2007-01	MCB944	272	0.0115 -	SFAM01.1	PPB	3/26/2023 23:05	SMPL
O2007-02	MCB945	299	0.0159 -	SFAM01.1	PPB	3/26/2023 23:07	SMPL

O2007-03	MCB949	1567	0.224 -	SFAM01.1	PPB	3/26/2023 23:09	SMPL
O2007-04	MCB952	575	0.0612 -	SFAM01.1	PPB	3/26/2023 23:12	SMPL
O2007-05	MCB954	304	0.0167 -	SFAM01.1	PPB	3/26/2023 23:14	SMPL
CCV005	CCV005	32311	5.2697 -	SFAM01.1	PPB	3/26/2023 23:16	SMPL
CCB005	CCB005	254	0.0085 -	SFAM01.1	PPB	3/26/2023 23:18	SMPL
O2007-06	MCB956	448	0.0404 -	SFAM01.1	PPB	3/26/2023 23:21	SMPL
O2007-07	MCB959	247	0.0074 -	SFAM01.1	PPB	3/26/2023 23:23	SMPL
O2007-08	MCB959D	279	0.0126 -	SFAM01.1	PPB	3/26/2023 23:25	SMPL
O2007-09	MCB959S	5332	0.8419 -	SFAM01.1	PPB	3/26/2023 23:28	SMPL
O2007-13	MCB966	300	0.0161 -	SFAM01.1	PPB	3/26/2023 23:30	SMPL
O2007-14	MCB967	294	0.0151 -	SFAM01.1	PPB	3/26/2023 23:32	SMPL
O2008-01	MCB947	258	0.0092 -	SFAM01.1	PPB	3/26/2023 23:34	SMPL
O2008-02	MCB950	454	0.0413 -	SFAM01.1	PPB	3/26/2023 23:37	SMPL
O2008-03	MCB951	266	0.0105 -	SFAM01.1	PPB	3/26/2023 23:39	SMPL
O2008-04	MCB953	292	0.0148 -	SFAM01.1	PPB	3/26/2023 23:41	SMPL
O2008-05	MCB955	283	0.0133 -	SFAM01.1	PPB	3/26/2023 23:43	SMPL
O2008-06	MCB957	287	0.0139 -	SFAM01.1	PPB	3/26/2023 23:46	SMPL
O2008-07	MCB960	268	0.0108 -	SFAM01.1	PPB	3/26/2023 23:48	SMPL
O2008-08	MCB960D	229	0.0044 -	SFAM01.1	PPB	3/26/2023 23:50	SMPL
O2008-09	MCB960S	5415	0.8555 -	SFAM01.1	PPB	3/26/2023 23:53	SMPL
PB151713BL	PBW	483	0.0461 -	SFAM01.1	PPB	3/26/2023 23:55	SMPL
O1886-01	MCB907	228	0.0042 -	SFAM01.1	PPB	3/26/2023 23:57	SMPL
O1886-02	MCB908	187	-0.0025 -	SFAM01.1	PPB	3/26/2023 23:59	SMPL
O1886-03	MCB911	560	0.0587 -	SFAM01.1	PPB	3/27/2023 0:02	SMPL
O1886-04	MCB913	293	0.0149 -	SFAM01.1	PPB	3/27/2023 0:04	SMPL
CCV006	CCV006	31936	5.2081 -	SFAM01.1	PPB	3/27/2023 0:06	SMPL
CCB006	CCB006	349	0.0241 -	SFAM01.1	PPB	3/27/2023 0:09	SMPL
O1886-05	MCB916	187	-0.0025 -	SFAM01.1	PPB	3/27/2023 0:11	SMPL
O1886-06	MCB919	324	0.02 -	SFAM01.1	PPB	3/27/2023 0:13	SMPL
O1886-07	MCB921	218	0.0026 -	SFAM01.1	PPB	3/27/2023 0:15	SMPL
O1886-08	MCB923	192	-0.0017 -	SFAM01.1	PPB	3/27/2023 0:18	SMPL
O1886-09	MCB925	155	-0.0077 -	SFAM01.1	PPB	3/27/2023 0:20	SMPL
O1886-10	MCB927	165	-0.0061 -	SFAM01.1	PPB	3/27/2023 0:22	SMPL
O1886-11	MCB929	807	0.0993 -	SFAM01.1	PPB	3/27/2023 0:25	SMPL
O1886-12	MCB931	241	0.0064 -	SFAM01.1	PPB	3/27/2023 0:27	SMPL
O1886-13	MCB933	154	-0.0079 -	SFAM01.1	PPB	3/27/2023 0:29	SMPL
O1886-14	MCB935	384	0.0298 -	SFAM01.1	PPB	3/27/2023 0:31	SMPL
O1886-15	MCB935D	374	0.0282 -	SFAM01.1	PPB	3/27/2023 0:34	SMPL
O1886-16	MCB935S	5080	0.8006 -	SFAM01.1	PPB	3/27/2023 0:36	SMPL
O1886-17	MCB937	161	-0.0067 -	SFAM01.1	PPB	3/27/2023 0:38	SMPL
O1886-18	MCB939	162	-0.0066 -	SFAM01.1	PPB	3/27/2023 0:40	SMPL
PB151714BL	PBW	371	0.0277 -	SFAM01.1	PPB	3/27/2023 0:43	SMPL
O1887-01	MCB909	180	-0.0036 -	SFAM01.1	PPB	3/27/2023 0:45	SMPL
O1887-02	MCB910	159	-0.0071 -	SFAM01.1	PPB	3/27/2023 0:47	SMPL
O1887-03	MCB912	184	-0.003 -	SFAM01.1	PPB	3/27/2023 0:50	SMPL
O1887-04	MCB914	176	-0.0043 -	SFAM01.1	PPB	3/27/2023 0:52	SMPL
O1887-05	MCB917	187	-0.0025 -	SFAM01.1	PPB	3/27/2023 0:54	SMPL

CCV007	CCV007	33061	5.3928 -	SFAM01.1	PPB	3/27/2023 0:56 SMPL
CCB007	CCB007	298	0.0157 -	SFAM01.1	PPB	3/27/2023 0:59 SMPL
O1887-06	MCB920	283	0.0133 -	SFAM01.1	PPB	3/27/2023 1:01 SMPL
O1887-07	MCB922	230	0.0046 -	SFAM01.1	PPB	3/27/2023 1:03 SMPL
O1887-08	MCB924	276	0.0121 -	SFAM01.1	PPB	3/27/2023 1:06 SMPL
O1887-09	MCB926	250	0.0079 -	SFAM01.1	PPB	3/27/2023 1:08 SMPL
O1887-10	MCB928	245	0.007 -	SFAM01.1	PPB	3/27/2023 1:10 SMPL
O1887-11	MCB930	253	0.0083 -	SFAM01.1	PPB	3/27/2023 1:13 SMPL
O1887-12	MCB932	313	0.0182 -	SFAM01.1	PPB	3/27/2023 1:15 SMPL
O1887-13	MCB934	285	0.0136 -	SFAM01.1	PPB	3/27/2023 1:17 SMPL
O1887-14	MCB936	330	0.021 -	SFAM01.1	PPB	3/27/2023 1:19 SMPL
O1887-15	MCB936D	288	0.0141 -	SFAM01.1	PPB	3/27/2023 1:22 SMPL
O1887-16	MCB936S	5004	0.7881 -	SFAM01.1	PPB	3/27/2023 1:24 SMPL
O1887-17	MCB938	233	0.0051 -	SFAM01.1	PPB	3/27/2023 1:26 SMPL
O1887-18	MCB940	259	0.0093 -	SFAM01.1	PPB	3/27/2023 1:29 SMPL
CCV008	CCV008	32259	5.2611 -	SFAM01.1	PPB	3/27/2023 1:31 SMPL
CCB008	CCB008	307	0.0172 -	SFAM01.1	PPB	3/27/2023 1:33 SMPL