



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB124690
Analysis Date: 03/27/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: Iwona
Extraction Date: 03/27/2023
Extraction IN Time: 08:30
Extraction OUT Time: 09:30
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB124690BL	LB124690BL	WATER	1.3	1000	100	3.0527	3.0527	3.02	3.0528	3.0528	0.0001	0.1
2	LB124690BS	LB124690BS	WATER	1.3	1000	100	3.1987	3.1987	3.01	3.2155	3.2155	0.0168	16.8
3	LB124690BSD	LB124690BSD	WATER	1.3	1000	100	3.0251	3.0251	3.03	3.0423	3.0423	0.0172	17.2
4	O1953-18	RR-TPH1L-WP	WATER	1.3	880	100	3.0736	3.0736	3.04	3.0928	3.0928	0.0192	21.82



Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2315
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0018</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1 :	<u>11:01</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>10:10</u>		
Bal Check Time:	<u>08:58</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1:	<u>11:38</u>
			Out Time1:	<u>11:00</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2 :	<u>12:56</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time2:	<u>12:15</u>		
Bal Check Time:	<u>13:35</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2:	<u>13:28</u>
			Out Time2:	<u>12:55</u>		

WORKLIST(Hardcopy Internal Chain)

1324690

WorkList Name : tph o2097 WorkList ID : 168516 Department : Wet-Chemistry Date : 03-27-2023 08:04:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/30/2023	Water	O1953-18	TPH	Conc H2SO4 to	CHEM02	QA Of	RR-TPH1L-WP	03/14/2023	1664A

Date/Time 03-27-23 08:15
 Raw Sample Received by: JDC (SM)
 Raw Sample Relinquished by: JDC (SM)

Date/Time 03-27-23 11:35
 Raw Sample Received by: JDC (SM)
 Raw Sample Relinquished by: JDC (SM)