

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB124718BLW	100	28/03/2023 10:35	0	3.28	3.27	104.1	7.03
LB124718BSW	100	28/03/2023 10:37	46.60	4.42	8.52	104.1	7.03
O2096-01	50	28/03/2023 10:43	64.60	4.32	6.26	104.1	7.03
O2096-01DUP	50	28/03/2023 10:47	64.28	4.30	6.25	104.1	7.03
O2096-03	50	28/03/2023 10:50	182.84	4.42	7.14	104.1	7.03
O2096-05	50	28/03/2023 10:54	183.48	4.48	7.05	104.1	7.03
O2096-07	50	28/03/2023 10:57	341.62	4.38	7.16	104.1	7.03
O1953-05	50	28/03/2023 11:02	328.22	4.48	8.25	104.1	7.03

03/28/2023

12

3/28/23