

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: SohilRun Number: LB124727Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB124727BL		1	24.2	0.82	0.83	03/28/2023	16:00
2	LB124727BS	1413	1	23.0	1370.00	1424.41	03/28/2023	16:05
3	O1953-07		1	24.7	552.00	555.18	03/28/2023	16:10

Note : Conductivity result is corrected to 25°C

Result = <u> (Reading)</u> 1 + 0.0191 (Temperature - 25)
--

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CONDUCTANCE-032823 **WorkList ID :** 168569 **Department :** Wet-Chemistry **Date :** 03-28-2023 13:58:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/30/2023	Water	O1953-07	Conductance	Cool 4 deg C	CHEM02	QA Of	PT-COND-WP LB124727	03/14/2023	SM2510 B
03/30/2023	Water	O1953-08	Conductance	Cool 4 deg C	CHEM02	QA Of	PT-COND-WP LB124728	03/14/2023	120.1
03/30/2023	Water	O1953-09	Conductance	Cool 4 deg C	CHEM02	QA Of	PT-COND-WP LB124729	03/14/2023	9050A

Date/Time 03/28/23 15:35
Raw Sample Received by: 12(SI)
Raw Sample Relinquished by: SI(OA office)

Date/Time 03/28/23 16:30
Raw Sample Received by: SI(OA office)
Raw Sample Relinquished by: 12(SI)